

Spatial Fuzzy C-Means Thresholding for Semiautomated Calculation of Percentage Lung Ventilated Volume From Hyperpolarized Gas and ^1H MRI

Paul J.C. Hughes, MEng,¹ Felix C. Horn, PhD,¹ Guilhem J. Collier, PhD,¹
Alberto Biancardi, PhD,^{1,2} Helen Marshall, PhD,¹ and Jim M. Wild, PhD^{1,2*}

Purpose: To develop an image-processing pipeline for semiautomated (SA) and reproducible analysis of hyperpolarized gas lung ventilation and proton anatomical magnetic resonance imaging (MRI) scan pairs. To compare results from the software for total lung volume (TLV), ventilated volume (VV), and percentage lung ventilated volume (%VV) calculation to the current manual “basic” method and a K-means segmentation method.

Materials and Methods: Six patients were imaged with hyperpolarized ^3He and same-breath lung ^1H MRI at 1.5T and six other patients were scanned with hyperpolarized ^{129}Xe and separate-breath ^1H MRI. One expert observer and two users with experience in lung image segmentation carried out the image analysis. Spearman (R), Intraclass (ICC) correlations, Bland–Altman limits of agreement (LOA), and Dice Similarity Coefficients (DSC) between output lung volumes were calculated.

Results: When comparing values of %VV, agreement between observers improved using the SA method (mean; $R = 0.984$, $\text{ICC} = 0.980$, $\text{LOA} = 7.5\%$) when compared to the basic method (mean; $R = 0.863$, $\text{ICC} = 0.873$, $\text{LOA} = 14.2\%$) nonsignificantly ($p_R = 0.25$, $p_{\text{ICC}} = 0.25$, and $p_{\text{LOA}} = 0.50$ respectively). DSC of VV and TLV masks significantly improved ($P < 0.01$) using the SA method (mean; $\text{DSC}_{\text{VV}} = 0.973$, $\text{DSC}_{\text{TLV}} = 0.980$) when compared to the basic method (mean; $\text{DSC}_{\text{VV}} = 0.947$, $\text{DSC}_{\text{TLV}} = 0.957$). K-means systematically overestimated %VV when compared to both basic (mean overestimation = 5.0%) and SA methods (mean overestimation = 9.7%), and had poor agreement with the other methods (mean ICC; K-means vs. basic = 0.685, K-means vs. SA = 0.740).

Conclusion: A semiautomated image processing software was developed that improves interobserver agreement and correlation of lung ventilation volume percentage when compared to the currently used basic method and provides more consistent segmentations than the K-means method.

Level of Evidence: 3

Technical Efficacy: Stage 2

J. MAGN. RESON. IMAGING 2017;00:000–000.

Respiratory diseases affect a large portion of the population, meaning development of sensitive imaging markers for diagnostic and prognostic assessment of lung disease is of growing importance. Ventilation-weighted hyperpolarized gas

(HP) and anatomical proton (^1H) lung magnetic resonance imaging (MRI) can be used for quantitative evaluation of lung function, including detection of early obstructive changes.¹ The most commonly used quantitative measures of

View this article online at wileyonlinelibrary.com. DOI: 10.1002/jmri.25804

Received Nov 10, 2016, Accepted for publication Jun 16, 2017.

This article presents independent research funded by the NIHR and MRC. The views expressed are those of the authors and not necessarily those of the NHS, the NIHR, the MRC, the Department of Health or GlaxoSmithKline.

*Address reprint requests to: J.M.W. Academic Unit of Radiology, Department of Infection Immunity and Cardiovascular Disease, University of Sheffield; Floor C, Royal Hallamshire Hospital, S10 2JF, Sheffield, South Yorkshire, UK. E-mail: j.m.wild@sheffield.ac.uk

From the ¹Academic Unit of Radiology, University of Sheffield, Sheffield, South Yorkshire, UK; and ²Insigneo Institute for in silico Medicine, Sheffield, South Yorkshire, UK

Additional supporting information may be found in the online version of this article

This is an open access article under the terms of the Creative Commons Attribution License, which permits use, distribution and reproduction in any medium, provided the original work is properly cited.

lung function derived from HP and ^1H scan pairs are lung ventilated volume percentage (%VV, the ratio of ventilated lung volume in HP images to total lung volume in ^1H images) and its counterpart the ventilation defect percentage (%VD = 100%-%VV).¹ Calculation of %VV, therefore, requires segmentation of both HP and ^1H image sets and, with manual lung segmentation taking on the order of 1–2 hours, depending on image resolution, this presents a time-consuming barrier for routine clinical application.

Development of methods to segment these scan pairs is thus necessary for quick, accurate, and reproducible quantitative analysis for clinical uptake of the methodology, while tackling technical challenges such as partial volume effects and motion artifacts from the heart.^{2,3}

Previous methods for HPG ventilation image segmentation have been based on manual intensity thresholding,¹ K-means clustering,^{4,5} multiple atlas labeling,⁶ and globally optimal graph cuts.⁷ He et al recently developed a method of characterizing the distribution of ventilation via linear binning,⁸ while Zha et al added an adaptive aspect of the K-means algorithm.⁹ Segmentation of ^1H anatomical scans has been proposed using a seeded region-growing algorithm,⁴ active contours within a closed homogeneous region,¹⁰ and a multiple atlas labeling approach.¹¹

Most of these segmentation techniques^{4,6–9,12} require little to no manual input, and the methods developed in Refs.⁴ and ⁸ also grade ventilation. However, the K-means segmentation method^{4,9} can fail, with low signal-to-noise ratio (SNR) images due to its binary clustering nature and inability to differentiate noise from lung tissue. A second potential disadvantage of the method developed previously⁴ is the Gaussian filtering of the ^1H images, which may lead to underestimation of the total lung volume. The performance of the method developed in Ref.⁷ also has limitations, with low SNR HPG images as the segmentation of the ^1H anatomical image is dependent on the HPG image itself. Although multiple atlas labeling^{6,11} may be the most automated approach, its complexity and need for suitable prior information are disadvantages.

Fuzzy C-means (FCM) clustering,^{12,13} which has been applied to HPG image segmentation,¹² is based on histogram information. The spatial Fuzzy C-means (SFCM) algorithm, developed previously,¹⁴ incorporates the use of spatial information into the calculation of the membership function and, unlike the standard FCM algorithm, it also allows use of neighboring pixel information when classifying a voxel, leading to a more robust segmentation in the presence of noise and partial volume effects.

The purpose of this work was to modify the SFCM method¹⁴ to segment both HPG and ^1H images of the lung and incorporate it into an image processing pipeline with high resilience to noise within a graphical user interface (GUI). Furthermore, in order to quantify the algorithm's

performance, the secondary aim was to compare the outputs of this novel semiautomated approach to that of the current basic segmentation and a K-means based method for %VV calculation.

MATERIALS AND METHODS

Patients scanned with ^3He were analyzed with local Research Ethics Committee approval, while patients scanned with ^{129}Xe gave informed consent as part of a separate research study.

Imaging

All imaging was carried out on a GE HDx 1.5T MR scanner (GE Healthcare, Milwaukee, WI). 3D ^1H anatomical (spoiled gradient echo [SPGR]) and HP ^3He ventilation-weighted (balanced steady-state free-precession [bSSFP]) images were acquired during the same breath-hold.^{2,15} 3D HP ^{129}Xe ventilation-weighted bSSFP images¹⁶ and 3D ^1H anatomical SPGR images were acquired in separate breaths. Sequence parameters are provided in the Supplementary Information.

Observers

Three observers (O1, O2, and O3) with 6, 1, and 5 years' experience in lung image segmentation, respectively, performed the analysis independently. O1, O2, and O3 analyzed ^3He scans and O2 and O3 analyzed ^{129}Xe scans, with both the basic method and the semiautomated method.

Participants

Six ^3He - ^1H scan pairs (^3He SNR range 47–72) were selected from a database of patients with respiratory conditions of various severities. Patients' ages ranged from 23–68 years (three male, three female) and forced expiratory volume in 1 second (FEV₁) (% predicted) ranged from 24–63%. One patient suffered from horseshoe lung, one patient had asthma, and four patients had chronic obstructive pulmonary disease (COPD).

In addition, six ^{129}Xe - ^1H scan pairs (^{129}Xe SNR range 18–34) from patients with lung cancer were analyzed to test the applicability of the method to HP ^{129}Xe ventilation images. Patient ages ranged from 62–76 (three male, three female), while FEV₁ (% predicted) ranged from 36–94%.

Image Analysis

The primary quantitative measure was %VV defined as ventilated volume (VV, from HPG ventilation-weighted images) divided by total lung volume (TLV, from ^1H anatomical images)¹:

$$\%VV = \frac{VV}{TLV} \times 100\% \quad (1)$$

The ventilated volume was masked by the total lung volume so that only voxels included in the TLV contributed to the calculated VV for all methods. ^{129}Xe and separate-breath ^1H acquisitions were coregistered using ANTs registration software.¹⁷

Basic Method

Segmentations were carried out using in-house software written in MatLab (MathWorks, Natick, MA). Thresholding was used to segment the ventilation images. ^1H images were segmented manually using the ventilation images overlaid on the ^1H images.




FIGURE 1: a: Workflow for image analysis; a(i) Raw ^1H anatomical image, single slice from 3D SPGR sequence, a(ii) ^1H mask single slice including airways, a(iii) ^1H mask single slice following removal of airways, a(iv) 3D representation of the ^1H mask, a(v) HP gas ventilation image single slice from 3D bSSFP sequence, a(vi) HP gas ventilation mask single slice including airways, a(vii) HP gas ventilation mask single slice following removal of airways, and a(viii) 3D representation of the ^3He mask. VV = (viii) and TLV = (iv). b: Segmentation method overview; b(i) original image, b(ii) image following application of bilateral filter, b(iii) cluster images, b(iv) selected cluster, and b(v) initial binary mask including airways.

Semiautomated Method

SUMMARY. A summary of the semiautomated image processing workflow is shown in Fig. 1. The GUI was implemented in Mat-Lab. Briefly, images were filtered using a bilateral filter with the filtering parameters tuned for ^3He , ^{129}Xe , and ^1H images. They were then clustered using the SFCM algorithm, using six clusters for ^1H anatomical images and four clusters for HP gas images. A single cluster was chosen that best represented the lung, then the membership values were thresholded to create a binary mask. Spatial and intensity membership weightings were both 1,¹⁴ however, the pipeline will modify the spatial weighting to 2 if image SNR drops below 20.

ALGORITHM DEVELOPMENT: FCM VS. SFCM. The FCM algorithm assigns N pixels to C clusters via Fuzzy memberships (μ).¹² These Fuzzy memberships do not take into account the spatial information, only intensity information. The SFCM algorithm, on the other hand, makes use of a window centered on each voxel of the image,¹⁴ which incorporates the membership information of

these voxels. This spatial information will then weight the membership function towards the correct cluster (with a specific weight being placed on the intensity and spatial memberships) only if the voxel was, for example, corrupted by noise, and would be incorrectly classified by the standard FCM method.

To decide whether to use FCM or SFCM as the clustering technique in this work, both FCM and SFCM segmentations were carried out on all ^3He and ^{129}Xe gas images and then assessed qualitatively (by visual inspection of features included/excluded by the algorithm) and quantitatively by comparing the unedited ventilated volume returned by the separate algorithms (the initial ventilation mask including airways).

Using the same number of clusters, the FCM method consistently included areas of low signal intensity deemed to be defects or noise via qualitative assessment. Additionally, there was a significant difference between the FCM and SFCM ventilated volumes for both ^3He and ^{129}Xe images ($P = 0.0312$ for both) (see Supplementary Information for detailed results; Fig. 2, for example, from one patient). Therefore, SFCM was used in the final semiautomated method.




FIGURE 2: (a) Original HPG image: unfiltered, (b) FCM mask overlaid on original image, and (c) SFCM mask overlaid on original image. The yellow circles point out some of the areas where FCM includes low-intensity values that are excluded in the SFCM method.

TABLE 1. Correlation, Bland-Altman and Intraclass Correlation Analysis of %VV for ^3He and ^{129}Xe Data

^3He data						
Parameter	O1 vs. O2	O1 vs. O3	O2 vs. O3	O1 vs. K-means	O2 vs. K-means	O3 vs. K-means
%VV Basic method						
R	0.909	0.859	0.819	0.905	0.850	0.894
ICC	0.850	0.920	0.850	0.726	0.660	0.670
Bias \pm LOA	-6.9% \pm 12.9%	-1.7% \pm 14.2%	-5.3% \pm 15.5%	-8.0% \pm 27.4%	-11.2% \pm 30.0%	-0.8% \pm 32.0%
%VV Semi-automated method						
R	0.988	0.981	0.978	0.901	0.905	0.926
ICC	0.990	0.970	0.980	0.702	0.730	0.788
Bias \pm LOA	-1.1% \pm 5.6%	-3.5% \pm 8.3%	2.4% \pm 8.7%	-11.2% \pm 29.6%	-10.1% \pm 28.4%	-7.7% \pm 26.5%
^{129}Xe data						
Parameter	O2 vs. O3	O2 vs. K-means	O3 vs. K-means			
%VV Basic method						
R	0.884	0.836	0.740			
ICC	0.579	0.610	0.356			
Bias \pm LOA	20.0% \pm 24.3%	-10.9% \pm 30.4%	-31.0% \pm 34.6%			
%VV Semiautomated method						
R	0.929	0.860	0.883			
ICC	0.905	0.589	0.660			
Bias \pm LOA	-0.9% \pm 19.1%	-13.3% \pm 33.8%	-12.4% \pm 28.5%			

FILTERING. To improve robustness and resilience against image noise and artifacts, a bilateral filter was used.¹⁸ The key property of this filter in the context of image segmentation is that edges are maintained due to the use of a nonlinear combination of range and domain filtering that weights pixel values depending on spatial and intensity similarity. Filter parameters that maintained ventilation defect integrity, smoothed artifact/noise, preserved edges, and ensured TLV was within an acceptable error margin ($\pm 5\%$) of manual segmentation were empirically determined by processing 12 sets of HP gas ventilation and ^1H anatomical image pairs. The training images (from patients with COPD and healthy volunteers) used for this optimization process were acquired in the same way, but are separate from the data used to evaluate algorithm performance.

Different filter values and mask thresholds were required for ^{129}Xe and ^3He images due to differing imaging resolutions and SNR. ^3He images (SNR range 25–67) filter had a window size of 3×3 and spatial and intensity standard deviation of 3 and 0.15, respectively, with a binary mask threshold (membership threshold) of 0.1. For ^{129}Xe images (SNR range 20–45) the filter intensity standard deviation was increased to 0.2 and the binary mask threshold decreased to 0.05. For ^1H images the intensity standard deviation was reduced to 0.1 and a binary mask threshold of 0.15 was used. All processing was carried out in-plane.

AUTOMATED/MANUAL EDITING. ^1H masks had data outside the lung region removed by a border-clearing algorithm (see

Supplementary Information for details). Main airways and vessels, and any regions of noise or tissue misclassified as lung volume, were removed manually from the masks using the software ITK-SNAP.¹⁹

K-means Method

All datasets were analyzed using a modified version of the method developed by Kirby et al.⁴ The size of the window used in the Gaussian filter for ^1H anatomical image segmentation was reduced from 15×15 to 3×3 and the standard deviation reduced to 0.01, the radius of the closing structuring element was reduced from 15 to 7, and data outside the lung region was removed by a border-clearing algorithm. No filtering of HP gas images was applied as per Ref. 4

Performance Evaluation

Performance analysis was carried out on %VV, VV, and TLV on a slice-by-slice basis. Intraclass correlation (ICC) using the two-way mixed model for absolute agreement, Bland–Altman analysis, and Spearman's or Pearson's correlation (R) were performed using GraphPad Prism (GraphPad Software, La Jolla, CA). Spatial comparisons were carried out on VV and TLV masks using the Dice Similarity Coefficient (DSC). *T*-tests and Wilcoxon signed rank tests were performed to assess statistical difference between analysis metrics.




FIGURE 3: a: Ventilation mask output by the basic, SFCM, and K-means method; a(i) Original image, a(ii) Basic mask overlaid on original image, a(iii) SFCM mask overlaid on original image, a(iv) K-means mask overlaid on original image. b: ^1H anatomical mask output by the basic, SFCM, and K-means method, b(i) original image, b(ii) basic mask overlaid on original image, b(iii) SFCM mask overlaid on original image, and b(iv) K-means mask overlaid on original image.

RESULTS

^3He

Table 1 shows the correlation, Bland–Altman limits of agreement (LOA) and ICC for %VV computed by different observers and methods. Correlation improved ($P = 0.25$) between observers when using the semiautomated method (%VV mean $R = 0.984$) when compared to the basic method (%VV mean $R = 0.863$), while mean ICC also increased ($P = 0.25$) from 0.873 using the basic method to 0.980 using the semiautomated method. LOA ($P = 0.50$) and %VV bias magnitude ($P = 0.25$) were reduced when using the semiautomated method (%VV mean LOA = 7.5%, mean $|\text{bias}| = 2.3\%$) when compared to the basic method (%VV mean LOA = 14.2%, mean $|\text{bias}| = 4.6\%$). These improvements were also seen in the VV and TLV measures. DSC significantly improved using the semiautomated method (VV mean DSC = 0.973, TLV mean DSC = 0.980) when compared to the basic method (VV mean DSC = 0.947, TLV mean DSC = 0.957) ($P < 0.01$ for both VV and TLV DSC).

The K-means method underestimated TLV when compared to both other methods. %VV was overestimated when compared to the basic (mean bias = 5.0%) and semiautomated (mean bias = 9.7%) methods (Fig. 3, for example). The Bland–Altman plot in Fig. 4a(iii) shows poor agreement of the K-means method %VV with the basic method for O2 and is representative of the pattern seen when comparing K-means with both the basic (%VV mean LOA = 29.8%) and semiautomated (%VV mean LOA = 28.2%) methods for all observers.

On average, the semiautomated method underestimated %VV by 4.6% when compared to the basic method carried out by the same observer, with a mean LOA of 19.7%. The semiautomated method reduced average segmentation time from 1 hour (basic) to 25 minutes.

^{129}Xe

Correlation, LOA, bias magnitude, and DSC improved between observers when using the semiautomated method when compared to the basic method (Fig. 4b(i,ii)). The K-means method underestimated TLV and overestimated %VV compared to the other methods to a greater extent than for the ^3He data. The semiautomated method underestimated %VV by 2.3% compared to the basic method for O2 (Fig. 4b(iv)) and overestimated %VV by 18.6% for O3. The mean LOA for %VV calculated by the same observer between the basic and semiautomated methods was 26.4%.

DISCUSSION

The semiautomated image processing workflow developed reduced interobserver variability, a problem in longitudinal imaging studies when multiple observers may be required. The use of the coregistered dual ^3He - ^1H image acquisition² in this work circumvents the need for image registration, which is commonly used in %VV analysis by other groups.^{4,8,20} However, image registration was required for the ^{129}Xe image analysis.

The semiautomated method results were more similar to those of the basic method than those obtained by the fully automated K-means method.⁴ Remaining differences in %VV values returned by the basic and semiautomated




FIGURE 4: (a) Bland-Altman analysis of ^3He %VV on a slice-by-slice basis, a(i) O2 v O3 basic, a(ii) O2 v O3 semi-automatic, a(iii) O2 basic v K-means, a(iv) O2 semi-automatic v K-means and (b) Bland-Altman analysis of ^{129}Xe %VV on a slice-by-slice basis, b(i) O2 v O3 basic, b(ii) O2vO3 semi-automatic, b(iii) O2 basic v K-means and b(iv) O2 basic v O2 semi-automatic.

method could be explained by the basic method including pixels that correspond to tissue on the ^1H image but where ^3He signal is also present, while the proposed method considers the edge of the lung from the ^1H image as the ground truth to exclude those pixels.

In the development phase of the semiautomated method, filtering followed by SFCM clustering was found to be robust to choice of imaging sequence and parameters. The technique returned similar masks of ventilated volume from HP gas images acquired with different sequences (SSFP/SPGR) and parameters (TE, TR) from the same subjects (see supplementary information). Semiautomatic segmentations of both the HP gas and proton images analyzed here were consistently of good quality, from images acquired with a range of fields of view (FOVs) (36–40 cm) and SNR (18–72) from patients with a variety of different diseases. Beyond removal of the main airways and vessels, little manual editing of the masks was required.

The K-means method underestimated TLV when compared to both other methods due to Gaussian filtering of the ^1H images, despite the filtering being substantially reduced in this work. This effect was exacerbated by the lower resolution of the ^1H images paired with the ^{129}Xe images. In addition, the original ventilation masks returned by the K-means method classified regions of noise as ventilated lung tissue. Both these factors lead to a systematic overestimation of %VV by the K-means method, which has implications for its use calculating outcome measures in clinical studies.

The limitations of this technical development study are the small numbers of patients analyzed as well as the reduced number of observers who segmented the ^{129}Xe images and the lack of comparison to other established techniques for %VV calculation.

In conclusion, the method presented here provides a robust and repeatable means of semiautomated lung MRI segmentation and was demonstrated on both ^3He and ^{129}Xe ventilation images. The method proposed improves interobserver agreement, and is easier and quicker to use than the current basic segmentation used to calculate lung ventilation volumes.

ACKNOWLEDGMENTS

Contract grant sponsor: National Institute of Health Research (NIHR); Contract grant sponsor: Medical Research Council (MRC); Contract grant sponsor: GlaxoSmithKline

The authors thank the University of Sheffield, GlaxoSmithKline, the NIHR and the MRC for funding this research. Paul J.C. Hughes: Co-funded PhD by GlaxoSmithKline and The University of Sheffield.

REFERENCES

- Woodhouse N, Wild JM, Paley MN, et al. Combined helium-3/proton magnetic resonance imaging measurement of ventilated lung volumes in smokers compared to never-smokers. *J Magn Reson Imaging* 2005; 21:365–369.
- Wild JM, Ajraoui S, Deppe MH, et al. Synchronous acquisition of hyperpolarised ^3He and ^1H MR images of the lungs—maximising mutual anatomical and functional information. *NMR Biomed* 2011;24: 130–134.
- Ivanovska T, Hegenscheid K, Laqua R, Gläser S, Ewert R, Völzke H. Lung segmentation of MR images: A review. In: *Visualization in Medicine and Life Sciences III*. New York: Springer; 2016. p 3–24.
- Kirby M, Heydarian M, Svenningsen S, et al. Hyperpolarized ^3He magnetic resonance functional imaging semiautomated segmentation. *Acad Radiol* 2012;19:141–152.
- Zha W, Kruger SJ, Cadman RV, et al. An adaptive K-means approach for assessment of ventilation defects in asthma and cystic fibrosis using hyperpolarized helium-3 MRI. In: *Proc 23rd Annual Meeting ISMRM*, Toronto; 2015.

6. Tustison NJ, Avants BB, Flors L, et al. Ventilation-based segmentation of the lungs using hyperpolarized ^3He MRI. *J Magn Reson Imaging* 2011;34:831–841.
7. Guo F, Yuan J, Rajchl M, et al. Globally optimal co-segmentation of three-dimensional pulmonary ^1H and hyperpolarized ^3He MRI with spatial consistency prior. *Med Image Anal* 2015;23:43–55.
8. He M, Driehuys B, Que LG, Huang Y-CT. Using hyperpolarized ^{129}Xe MRI to quantify the pulmonary ventilation distribution. *Acad Radiol* 2016;23:1521–1531.
9. Zha W, Niles DJ, Kruger SJ, et al. Semiautomated ventilation defect quantification in exercise-induced bronchoconstriction using hyperpolarized helium-3 magnetic resonance imaging: a repeatability study. *Acad Radiol* 2016;23:1104–1114.
10. Ray N, Acton ST, Altes T, De Lange EE, Brookeman JR. Merging parametric active contours within homogeneous image regions for MRI-based lung segmentation. *IEEE Trans Med Imaging* 2003;22:189–199.
11. Tustison NJ, Qing K, Wang C, Altes TA, Mugler JP. Atlas-based estimation of lung and lobar anatomy in proton MRI. *Magn Reson Med* 2015 [Epub ahead of print].
12. Lui JK, LaPrad AS, Parameswaran H, Sun Y, Albert MS, Lutchen KR. Semiautomatic segmentation of ventilated airspaces in healthy and asthmatic subjects using hyperpolarized MRI. *Computat Math Methods Med* 2013;2013.
13. Bezdek JC, Ehrlich R, Full W. FCM: The fuzzy c-means clustering algorithm. *Comput Geosci* 1984;10:191–203.
14. Chuang K-S, Tzeng H-L, Chen S, Wu J, Chen T-J. Fuzzy c-means clustering with spatial information for image segmentation. *Comput Med Imaging Graph* 2006;30:9–15.
15. Horn F, Tahir B, Stewart N, et al. Lung ventilation volumetry with same-breath acquisition of hyperpolarized gas and proton MRI. *NMR Biomed* 2014;27:1461–1467.
16. Stewart NJ, Norquay G, Griffiths PD, Wild JM. Feasibility of human lung ventilation imaging using highly polarized naturally abundant xenon and optimized three-dimensional steady-state free precession. *Magn Reson Med* 2015 [Epub ahead of print].
17. Avants BB, Tustison N, Song G. Advanced normalization tools (ANTS). *Insight J* 2009.
18. Tomasi C, Manduchi R. Bilateral filtering for gray and color images. *Computer Vision, IEEE 1998 Sixth International Conference*; 1998. p 839–846.
19. Yushkevich PA, Piven J, Hazlett HC, et al. User-guided 3D active contour segmentation of anatomical structures: significantly improved efficiency and reliability. *NeuroImage* 2006;31:1116–1128.
20. He M, Kaushik SS, Robertson SH, et al. Extending semiautomatic ventilation defect analysis for hyperpolarized ^{129}Xe ventilation MRI. *Acad Radiol* 2014;21:1530–1541.