



Deposited via The University of Leeds.

White Rose Research Online URL for this paper:

<https://eprints.whiterose.ac.uk/id/eprint/167505/>

Version: Accepted Version

Article:

Zhong, B, Cao, J, McDaid, A et al. (2020) Synchronous Position and Compliance Regulation on a Bi-Joint Gait Exoskeleton Driven by Pneumatic Muscles. IEEE Transactions on Automation Science and Engineering, 17 (4). pp. 2162-2166. ISSN: 1545-5955

<https://doi.org/10.1109/tase.2020.2992890>

© 20xx IEEE. Personal use of this material is permitted. Permission from IEEE must be obtained for all other uses, in any current or future media, including reprinting/republishing this material for advertising or promotional purposes, creating new collective works, for resale or redistribution to servers or lists, or reuse of any copyrighted component of this work in other works.

Reuse

Items deposited in White Rose Research Online are protected by copyright, with all rights reserved unless indicated otherwise. They may be downloaded and/or printed for private study, or other acts as permitted by national copyright laws. The publisher or other rights holders may allow further reproduction and re-use of the full text version. This is indicated by the licence information on the White Rose Research Online record for the item.

Takedown

If you consider content in White Rose Research Online to be in breach of UK law, please notify us by emailing eprints@whiterose.ac.uk including the URL of the record and the reason for the withdrawal request.

Synchronous Position and Compliance Regulation on a Bi-Joint Gait Exoskeleton Driven by Pneumatic Muscles

Bin Zhong[#], Jinghui Cao[#], Andrew McDaid, *IEEE Member*, Sheng Quan Xie, *Senior IEEE Member*
Mingming Zhang, *IEEE Member*

Abstract—A previously-developed pneumatic muscles (PM) actuated gait exoskeleton (with only knee joint) has been demonstrated in achieving appropriate actuation torque, range of motion (ROM), and control bandwidth for task-specific gait training. While the adopted multi-input-multi-output (MIMO) sliding mode (SM) strategy has preliminarily implemented simultaneous control of the exoskeleton’s angular trajectory and compliance, its efficacy with human users during gait cycles has not been investigated. This short paper presents an improved Bi-joint Gait Rehabilitation Exoskeleton (BiGREX) with integrated human hip and knee joints. Results with 12 healthy subjects demonstrated that the system’s compliance can be effectively adjusted while guiding the subjects walking in predefined trajectories.

Note to Practitioners—This paper was motivated by achieving compliant interaction between PM-actuated exoskeletons and human when conducting task-specific gait training. Due to the intrinsic nonlinearity of PM, it’s challenging to establish a mathematical model to precisely predict real-time compliance of the powered joints. This paper suggests a new strategy that adopts the average pressure of flexor and extensor PMs as the feedback to synchronously realize the joint position control and compliance regulation. A novel experimental approach was adopted to validate the system capability on adjusting the compliance from human users’ perception. This paper provides a new insight between the controlled PM pressure and desired joint-compliance, which would be essential for the future design of PM-actuated exoskeletons.

Index Terms—MIMO control, compliant interaction, PM lower limb exoskeleton.

I. INTRODUCTION

Cao et al. [1] developed a MIMO SM controller on a knee joint mechanism driven by antagonistic PMs to simultaneously track the mechanism’s angular trajectory and average antagonistic PM pressure. The robotic behaviors are comparable to algorithms of variable impedance reported on motor-driven gait rehabilitation devices [2]. However, the MIMO SM based compliance adaptation performance has not been investigated during fully actuated human gait cycles.

This paper is an extended study based on the MIMO SM control strategy [1] with two major contributions. One is the improved BiGREX system integrated with both human hip and knee joints to mimic natural gait cycles, and the other is to

experimentally investigate if the MIMO SM technique can vary the robotic assistance provided to human users while guiding them to walk along a reference gait trajectory. To the best authors’ knowledge, it is the first attempt to implement variable compliance on a fully PM actuated gait exoskeleton with systematic experimental validation.

II. SYSTEM AND MODELING

Fig. 1 presents the configuration of BiGREX, control diagram of the MIMO SM controller and its use with BiGREX system on a human user. A myRIO platform by National Instruments is employed for data acquisition (at 1000 Hz) and real-time control processing (at 100 Hz). A custom printed circuit board was also designed for physical hardware interfacing the myRIO. The BiGREX system mainly consists of a trunk mechanism with three unactuated degrees of freedoms (DoFs) and a bi-joint mechanism with two powered joints. Detailed description of single joint kinetics, load dynamics of single joint and the dynamics of chosen FESTO PM can refer to [1]. The modeling of stiffness S which can be defined as $\Delta\tau/\Delta\theta$ is required to vary the compliance when controlling the position. The theoretical stiffness can be derived from different models, like dynamic model from our previous work [1, 3] or static force models proposed in [4-6], etc.

The first approach to calculate joint compliance is using the PM’s dynamic force model. Based on our previous work, the output torque of the bi-directional PMs can be expressed as equation (1) [1]:

$$\tau = \tau_F - \tau_E = 2r[F(P_F) + B(P_F)\dot{x}_F + K(P_F)x_F - (F(P_E) + B(P_E)\dot{x}_E + K(P_E)x_E)] \quad (1)$$

where $F(P)$, $B(P)$ and $K(P)$ denote the pressure-dependent force, damping and spring parameters based on the classic dynamic model of pneumatic muscle in [7]; x denotes the contracting length of extensor and flexor PMs; r denotes pulley radius of an actuated joint; and parameters with subscripts F and E are corresponding to the flexion and extension motion, respectively. Nevertheless, the joint compliance is dominated by the nominal pressure of the PMs [8]. Hence, by extracting the spring parameter forms in (1), the output torque of the joint’s spring force terms can be derived as:

$$\begin{aligned} \tau &= \tau_F - \tau_E = 2r[K(P_F)x_F - K(P_E)x_E] \\ &= 2r[(K_0 + K_1(P_{avg} + \Delta P_F)r\theta) \\ &\quad - (K_0 + K_1(P_{avg} - \Delta P_E)r(-\theta))] \end{aligned} \quad (2)$$

here the P_{avg} denotes the desired average antagonistic PM pressure; ΔP_F and ΔP_E denote the difference of the nominal pressure and P_{avg} of the flexor and extensor; K_0 and K_1 were determined experimentally; θ denotes the joint angle. Hence, the defined joint compliance \mathbb{C} can be expressed as:

[#]Co-authors Bin Zhong and Jinghui Cao contributed equally to this work. Bin Zhong (11955001@mail.sustech.edu.cn), Mingming Zhang* (zhangmm@sustech.edu.cn) are with the Department of Biomedical Engineering at the Southern University of Science and Technology, China.

Jinghui Cao (jcao027@aucklanduni.ac.nz), and Andrew McDaid (andrew.mcdaid@auckland.ac.nz) are with the Department of Mechanical Engineering at the University of Auckland, New Zealand.

Shengquan Xie (s.q.xie@leeds.ac.uk) is with the School of Electronic and Electrical Engineering at the University of Leeds, Leeds, UK.




Fig. 1 The control diagram of the MIMO SM controller and its use with the BiGREX system on a human user. And the Bi-joint mechanism is illustrated in the right, both the thigh (in red) and shank (in blue) segments are powered by four pneumatic muscles.

$$\mathbb{C} = \frac{1}{s} = \frac{\theta}{\tau} = \frac{1}{2r^2[(2K_0 + 2K_1 P_{avg} + K_1(\Delta P_F - \Delta P_E)]} \quad (3)$$

When the MIMO SM controller can perfectly track the desired average antagonistic PM pressure ($\Delta P_F = \Delta P_E$), the joint stiffness is linearly dependent on the set average pressure. With both K_0 and K_1 being set as positive constants, the compliance \mathbb{C} decreases as P_{avg} increases.

The second approach is applying a static force model by Chou and Hannaford [4], which expresses the PM instant contractile force F as:

$$F = \left(\frac{3(l_0 - x)^2 - b^2}{4\pi n^2} \right) P \quad (4)$$

where l_0 is the original length of the PM; b and n are constants related to the geometry of the PM; P denotes the instant pressure of antagonistic PMs, and equation (4) is appropriate for both flexor and extensor PM. Through substituting F into equation (1) and extracting only the spring parameter forms, the desired compliance, when the average pressure can be tracked as expected ($\Delta P_F = \Delta P_E$), can be calculated as:

$$\mathbb{C} = \frac{\pi n^2}{3r^2[-2l_0 + r(\theta_{E0} - \theta_{F0})P_{avg} + (x_F - x_E)\Delta P]} \quad (5)$$

where θ_{E0} and θ_{F0} denote joint angular positions when the extensor and flexor PMs are at their original lengths.

It's obviously demonstrated by equations (3) and (5) that the relationship between compliance and pressure is inversely proportional. However, we can also notice that the joint compliance cannot be precisely derived from the PM models which have been proved to be practical in certain cases, because (1) both models are simplified due to the nonlinearity of the PM; (2) different models present varied results. Note that, no extra mechanism was implemented to measure the real-time compliance magnitude and feedback it for control. Hence, it would be essential to conduct experiments to prove the

feasibility of the proposed MIMO SM algorithm with close-loop average PM pressure control model during human gait. The performance of adaptive compliance control will be presented in Section III-C.

III. EXPERIMENTS AND RESULTS

Comparing to the preliminary experiments on knee joint in [1], three sets of experiments were designed in this study with natural human walking gait aiming to comprehensively evaluate the effectiveness of the MIMO SM control with the compliance adaptable BiGREX system, results will be presented in section A to C.

Ethical approval for experiments was granted by the University of Auckland Human Participants Ethics Committee (014970). Written informed consent had been obtained from all the participants prior to conducting any experiments.

A. Technical Validation of BiGREX's Control System

This experiment aims to validate the control system's ability to track the desired gait trajectories and average pressures of both actuated joints simultaneously. The experiment was performed without human subject attached. During the experiment, both the hip and knee joints of BiGREX was programmed to track reference gait trajectories in joint space. The desired average antagonistic PM pressures of both actuated joints were set to 270 KPa, a moderate level of system compliance, with the experimental results shown in Fig.2.

It can be observed from Fig. 2 that the proposed MIMO SM control system is able to track a gait trajectory while maintaining the average PM pressures of both actuated joints. In general, the hip joint performed better in both trajectory (RMSE being 0.022 rad for hip and 0.125 rad for knee) and average PM pressure tracking (RMSE being 0.746 KPa for hip and 16 KPa for knee). This could be due to that the desired knee joint trajectory has higher ROM and was more challenging for the controller to track. Considering the safety and nature of




Fig. 2 The joint space trajectory and average PM pressure tracking performance for both actuated joints of GAREX. **Top**: the trajectory tracking performance of the hip joint. **Middle**: the trajectory tracking performance of the knee joint. **Bottom**: the average pressure control performance of both the hip and knee joints. RMSE: Root mean square error.

robotic gait rehabilitation, angular trajectory tracking needs to be prioritized among the two control objectives. This was implemented by the tuning parameters that let angular trajectory tracking take the dominant role in the overall control actions. As a result, the pressure tracking performance was slightly sacrificed, especially when the desired rate of position change is significant. This is the reason that pressure tracking performance is better on the hip joint.

B. Task-specific Gait Training Pilot

The automated treadmill-based gait training performance of BiGREX with human users will be investigated in this section. One healthy subject with the medium build (male, 172 cm, 68 kg) participated in the validation experiment of the robotic gait training. The exoskeleton was adjusted according to the participant's anthropometric data. During the experiment, programmed with the MIMO SM controller, BiGREX guided the subject to walk in a predefined gait trajectory on the treadmill with a speed of 1.5 km/h. The participant was asked to relax his leg strapped to the exoskeleton and let BiGREX in charge. The participant also adapted unactuated leg's movements to achieve stable walking on the treadmill. Four two-minute experiments were conducted and in between two experiments, there was a break of two minutes. In each experiment, the average PM pressures of the hip and knee joints were controlled to the same magnitude. The desired average antagonistic PM pressures were in turn 160, 240, 320 and 400 KPa for four experiments. Combining the dynamic algorithm from average antagonistic PM pressure to joint compliance

described in [9], the ideal joint compliance under these four experimental conditions were 0.0158, 0.0152, 0.0147 and 0.0143 rad/Nm, respectively.

During the training process, this participant was able to walk smoothly on the treadmill with the guidance from BiGREX with different joint compliance levels. His gait trajectories of the two actuated joints are presented in Fig. 3, where gait data over the 2-minute experimental intervals were averaged, and the trajectory tracking performance was evaluated with the RMSE value. It is obvious that the RMSE had a trend of decrease as the controlled average PM pressure increased, indicating that BiGREX became less compliant and gave the subject less freedom around the desired trajectory, and providing a statistical basis for the optimal predefinition of joint compliance to a specific user.

C. Validation of Controllable Compliance

As shown in Fig.1, the average pressures of PMs were adopted in the closed-loop control algorithm. To further validate whether the proposed controller is capable of tracking predefined gait trajectory and simultaneously adjusting the actuation compliance to modify the assistance provided, the third set of experiments was conducted. A total of 12 (10 males and 2 females) subjects of ages between 23 and 31 were involved, with height ranging from 164 to 188 cm and weight from 49 to 100 kg. Before a participant starting experiments, the concept of compliance had been explained.

Each participant was required to conduct four experiments. Each experiment lasted approximated three minutes in between the two experiments there was a break of two minutes. Similar to the experiment described in Section III-B, all participants were assisted by the robot to walk on the treadmill at the speed of 1.5 km/h. The average PM pressures of the hip and knee joints were controlled at 160, 240, 320 and 400 KPa.

Each participant was required to experience four experiments with different pressures in a randomized order, and he/she was not informed of the controlled average pressure. Each of them was asked to relax and let the BiGREX guide the leg movement for the first half of an experiment. In the second half, he/she was asked to actively walk with the assistance from the exoskeleton and/or try to walk in his/her preferred gait pattern disregarding the assistance from the exoskeleton, so that the subject would have the impression of the compliance of the exoskeleton. After completing the first two and three trials, the subject was asked to rank the compliance levels of the experiment he/she had completed. After finishing all four trials, the subject was able to rank the compliance level of each experiment with a discrete score of 1, 2, 3 or 4, with 1 being most compliant and 4 being least compliant. No matter what result was given after the four experiments, the researcher asked the subject if he/she wished to re-conduct any of the experiments to help him/her refresh memory or eliminate any uncertainty. Each participant could request up to two experiments to be replayed. For each controlled average pressure, the means and variances of the compliance scores rated by the participants were calculated. Such results are summarized in the first two rows of Table I. One-way Analysis of Variance (ANOVA) was thought to be suitable for examining if there is any statistical difference between the means of compliance scores. Before that, the *Levene* test was conducted to investigate the equality of




Fig. 3 The comparison between the desired and actual gait trajectories of the two actuated joints during the validation experiments. The trajectories have been normalized to one gait cycle. The red lines are the predefined reference gait trajectory. The blue lines represent the average gait trajectories over the experimental period. The shaded area stands for the standard deviations of the average trajectories over the recorded gait cycles. Subplots (a, c, e, g) are the trajectories of the knee joint and subplots (b, d, f, h) are for the hip joint. The subplots in the top, second, third and bottom rows are for experiments with desired average antagonistic PM pressures 160, 240, 320 and 400 KPa respectively.

variance of the experimental data. The test's p -value of 0.032 rejects the null hypothesis that all four groups have equal variance. A commonly used logarithm transformation of the original experimental data has been performed and the means and variances of the transformed data are listed in the third and fourth rows of Table I. The *Levene* test of transformed results gave a p -value of 0.462, which means the assumption of equal variance is not violated.

The one-way ANOVA analysis is then performed on the transformed data and its p -value of 0.000 suggests a very strong evidence to reject the null hypothesis that all the samples have equal means. Multiple comparisons via Tukey's test was also conducted to investigate whether there are statistical differences between all the combinations of two different groups. The analytical results again expressed by p -values are exhaustively listed in the second half of TABLE I. As seen from the table, there is highly significant difference (p -values: 0.000) in means between groups whose average pressure levels are not adjacent. For adjacent groups, there is very strong evidence for the difference in means between 160 and 240 KPa. There is strong evidence for the difference in means between 240 and 400 KPa

groups. However, there is not significant statistical difference in means between the 320 and 400 KPa groups.

The above statistical analysis further proves that there is an inversely proportional relationship between the controlled average pressure of the PM actuation system and the

TABLE I
RESULTS OF COMPLIANCE SCORES AND ANOVA ANALYSIS

| | Controlled Average PM Pressures | | | |
|----------------------|---------------------------------|--------|--------|--------|
| | 160KPa | 240KPa | 320KPa | 400KPa |
| Mean | 1.083 | 2.333 | 3.000 | 3.583 |
| Variance | 0.083 | 0.606 | 0.545 | 0.447 |
| Mean Transformed | 0.025 | 0.320 | 0.464 | 0.571 |
| Variance Transformed | 0.008 | 0.016 | 0.013 | 0.003 |
| 160 KPa | | 0.000 | 0.000 | 0.000 |
| 240 KPa | | | 0.005 | 0.000 |
| 320 KPa | | | | 0.056 |

Note: The upper half of the table shows a bunch of statistic values of tested compliance. The bottom half shows all the possible p -values between groups via the Tukey's test. Gray colored grids denote nonsignificant difference.

exoskeleton compliance. However, the relationship is not linear. For the same difference interval in average pressure, the change in compliance is more significant when the average pressure is lower. Understanding such a trend is important since further development of the BiGREX system may require the automatic adaptation of the compliance level based on the patient's performance or ability assessment during rehabilitation training.

IV. REMARKS AND CONCLUSIONS

A unilateral robotic gait rehabilitation system- BiGREX actuated by PMs has been well evaluated with human users. Only a few PM driven gait rehabilitation exoskeletons have been reported and to our knowledge, none of them have conducted detailed compliance analysis to verify the system's ability to provide task-specific gait training. BiGREX was designed with future clinical applications in mind and possibilities of combining with extra ankle rehabilitation mechanisms to form a full limb exoskeleton [5, 10]. The system is capable to provide sufficient torque for the knee and hip joint rotations in the sagittal planes to facilitate gait rehabilitation for a patient weighing up to 100 kg.

It's worth noting that task-specific robotic gait training imposes ROM and bandwidth requirements to the system development. While the previous PM driven robotic rehabilitation system usually ignored the intrinsic slow pressure dynamics of PMs by using pressure regulating valves, which could potentially affect the controlled bandwidth [11]. For BiGREX, the use of the analogue valves enabled the researchers to obtain the entire system model that encompasses all major uncertainties.

Based on such a model, the MIMO SM controller was thus developed and tuned to deliver robust control performance. Currently, BiGREX can comfortably guide treadmill based gait training at a speed of 1.5 km/h, which is similar to the slower speed adopted for clinical trials on motor-driven gait rehabilitation exoskeleton [12]. At the moment, the controlled bandwidth is limited by the pneumatic supply system and the mechanical design. To further increase the bandwidth the following changes could be made: 1) increasing current pneumatic supply pressure (6 bar); 2) using larger diameter tubes and valves; 3) optimizing the exoskeleton design to further reduce its weight.

The means of using average pressure to represent actuation compliance is not new. Due to the system simplification with pressure regulating valves, the joint space trajectory and average PM pressure controller are separated and the average PM pressure was adjusted through an open-loop controller. The actuation compliance is again calculated from the average PM pressure using the force dynamics model of the PM. There were also no dedicated experiments to investigate if such a compliance adjusting approach could vary the extent of guidance during robotic gait rehabilitation.

BiGREX's MIMO SM control system considers PM pressures as state variables and allows more direct manipulation of the average PM pressure compared to [13-15]. A novel experimental approach was adopted to validate the system capability on adjusting the compliance from participant's perception, which guarantees a safe and compliant human-robot interaction so as to enhance the patients' trust in robots. It was

demonstrated in another of our researches that compliant interaction and trust in robots are key factors that contribute to enhanced rehabilitation performance [16]. Moreover, statistical analysis indicated the system is able to provide different levels of assistance via changing the compliance when providing task-specific gait training.

REFERENCES

- [1] J. Cao, S. Q. Xie, and R. Das, "MIMO sliding mode controller for gait exoskeleton driven by pneumatic muscles," *IEEE Transactions on Control Systems Technology*, vol. 26, no. 1, pp. 274-281, 2018.
- [2] B. Vanderborgh et al., "Variable impedance actuators: A review," *Robotics and Autonomous Systems*, vol. 61, no. 12, pp. 1601-1614, 2013, doi: 10.1016/j.robot.2013.06.009.
- [3] J. Cao, S. Q. Xie, M. Zhang, and R. Das, "A new dynamic modelling algorithm for pneumatic muscle actuators," in *International Conference on Intelligent Robotics and Applications*, 2014: Springer, pp. 432-440.
- [4] C.-P. Chou and B. Hannaford, "Measurement and modeling of McKibben pneumatic artificial muscles," *IEEE Transactions on robotics and automation*, vol. 12, no. 1, pp. 90-102, 1996.
- [5] M. Zhang, A. McDaid, A. J. Veale, Y. Peng, and S. Q. Xie, "Adaptive Trajectory Tracking Control of a Parallel Ankle Rehabilitation Robot With Joint-Space Force Distribution," *IEEE Access*, vol. 7, pp. 85812-85820, 2019, doi: 10.1109/access.2019.2925182.
- [6] J. Sarosi, "New approximation algorithm for the force of Fluidic Muscles," presented at the 2012 7th IEEE International Symposium on Applied Computational Intelligence and Informatics (SACI), 2012.
- [7] D. Reynolds, D. Repperger, C. Phillips, and G. Bandry, "Modeling the dynamic characteristics of pneumatic muscle," *Annals of biomedical engineering*, vol. 31, no. 3, pp. 310-317, 2003.
- [8] S. Hussain, S. Q. Xie, and P. K. Jamwal, "Robust nonlinear control of an intrinsically compliant robotic gait training orthosis," *IEEE Transactions on Systems, Man, and Cybernetics: Systems*, vol. 43, no. 3, pp. 655-665, 2012.
- [9] S. Xie, M. Zhang, and W. Meng, *Soft Robots for Healthcare Applications: Design, Modeling, and Control*. Institution of Engineering and Technology, 2017.
- [10] M. Zhang, J. Cao, G. Zhu, Q. Miao, X. Zeng, and S. Q. Xie, "Reconfigurable workspace and torque capacity of a compliant ankle rehabilitation robot (CARR)," *Robotics and Autonomous Systems*, vol. 98, pp. 213-221, 2017, doi: 10.1016/j.robot.2017.06.006.
- [11] M. Zhang et al., "Adaptive Patient-Cooperative Control of a Compliant Ankle Rehabilitation Robot (CARR) with Enhanced Training Safety," *Ieee T Ind Electron*, vol. 65, no. 2, pp. 1398 - 1407, 2017.
- [12] K. P. Westlake and C. Patten, "Pilot study of Lokomat versus manual-assisted treadmill training for locomotor recovery post-stroke," *J Neuroeng Rehabil*, vol. 6, p. 18, 2009.
- [13] T.-Y. Choi, B.-S. Choi, and K.-H. Seo, "Position and compliance control of a pneumatic muscle actuated manipulator for enhanced safety," *Control Systems Technology, IEEE Transactions on*, vol. 19, no. 4, pp. 832-842, 2011.
- [14] S. Hussain, S. Q. Xie, and P. K. Jamwal, "Robust Nonlinear Control of an Intrinsically Compliant Robotic Gait Training Orthosis," *Systems, Man, and Cybernetics: Systems, IEEE Transactions on*, vol. 43, no. 3, pp. 655-665, 2013, doi: 10.1109/TSMCA.2012.2207111.
- [15] Q.-T. Dao and S.-i. Yamamoto, "Assist-as-needed control of a robotic orthosis actuated by pneumatic artificial muscle for gait rehabilitation," *Applied Sciences*, vol. 8, no. 4, p. 499, 2018.
- [16] B. Zhong, W. Niu, E. Broadbent, A. McDaid, T. M. C. Lee, and M. Zhang, "Bringing Psychological Strategies to Robot-Assisted Physiotherapy for Enhanced Treatment Efficacy," *Front Neurosci*, vol. 13, p. 984, 2019, doi: 10.3389/fnins.2019.00984.



Bin Zhong received the B.E. and the M.E. degrees in “Mechanical Engineering and Automation” and “Instrument Science and Technology” from the Department of Precision Machinery and Precision Instrumentation (PMPI), University of Science and Technology of China (USTC), Hefei, China, in 2015 and 2018. He is currently a joint Ph.D. at the Southern University of Science and Technology and the National University of

Singapore. His research interests include lower limb rehabilitation robots, serial elastic actuator and biomechanics.



Jinghui Cao received the B.E. with a degree in Mechatronics Engineering with First Class Honors from the University of Auckland, New Zealand in 2013. Since March 2013, he has been working towards a Ph.D. degree in the Department of Mechanical Engineering, University of Auckland. During this research, he published over 8 papers and 2 book chapters. His research interests include mechatronics, rehabilitation robotics, and motion control.



Andrew McDaid received the B.E. degree with First Class Honor and Ph.D. from the University of Auckland in 2008 and 2012. He is a senior lecturer at the University of Auckland, New Zealand. He has published more than 60 research papers and has three provisional patents. His main research interests include design and control medical devices, functional electrical stimulation of neuro-

prosthetics, smart materials and soft robotics, physiological control (EEG, HD-EMG), gamification and developing intelligent medical robotics for people who have suffered injury, stroke or have disabilities.



Shengquan Xie received the Ph.D. degrees from Huazhong University of Science and Technology, Wuhan, China, and the University of Canterbury, Christchurch, New Zealand, in 1998 and 2002, respectively. From 2003 to 2016, he worked in the University of Auckland, New Zealand, where he chaired the research group of Bio-mechatronics. Since 2017, he has joined the University of Leeds.

He has published 5 books, 15 book chapters,

and more than 280 academic papers. His research interests include medical and rehabilitation robots and advanced robot control. Prof. Xie was elected a Fellow of The Institution of Professional Engineers New Zealand in 2016. He has also served as a Technical Editor of the IEEE/ASME Transactions on Mechatronics.



Mingming Zhang received the M.Eng. degree in mechatronics from Chongqing University, China, in 2012, and Ph.D. degree in mechanical engineering from the University of Auckland, Auckland, New Zealand, in 2016. Since 2015, he has been working as a research fellow and/or visiting research fellow in the Department of Auckland. His research includes mechatronics, medical robotics, biomechanics, and advanced control techniques. He has authored over 30 academic

papers and 1 book. He also holds over 20 patents. Dr. Zhang has served as the invited reviewer for many high-quality international journals,

including IEEE Transactions on Mechatronics, IEEE Transactions on Biomedical and Engineering, and IEEE Transactions on Industrial Electronics.