



This is a repository copy of *The use of biofeedback for gait retraining: a mapping review*.

White Rose Research Online URL for this paper:
<https://eprints.whiterose.ac.uk/181606/>

Version: Accepted Version

Article:

van Gelder, L.M.A. orcid.org/0000-0002-4124-2052, Barnes, A., Wheat, J.S. et al. (1 more author) (2018) The use of biofeedback for gait retraining: a mapping review. *Clinical Biomechanics*, 59. pp. 159-166. ISSN 0268-0033

<https://doi.org/10.1016/j.clinbiomech.2018.09.020>

Article available under the terms of the CC-BY-NC-ND licence
(<https://creativecommons.org/licenses/by-nc-nd/4.0/>).

Reuse

This article is distributed under the terms of the Creative Commons Attribution-NonCommercial-NoDerivs (CC BY-NC-ND) licence. This licence only allows you to download this work and share it with others as long as you credit the authors, but you can't change the article in any way or use it commercially. More information and the full terms of the licence here: <https://creativecommons.org/licenses/>

Takedown

If you consider content in White Rose Research Online to be in breach of UK law, please notify us by emailing eprints@whiterose.ac.uk including the URL of the record and the reason for the withdrawal request.



eprints@whiterose.ac.uk
<https://eprints.whiterose.ac.uk/>

Title:

The use of biofeedback for gait retraining: A mapping review

Authors:

Linda M. A. van Gelder ^[a], Andrew Barnes^[b], Jonathan S. Wheat^[b], Ben W. Heller^[a]

Affiliations:

^[a] Sheffield Hallam University, Faculty of Health and Wellbeing, Centre for Sports

Engineering Research, 11 Broomgrove Road, Sheffield S10 2LX , United Kingdom

^[b] Sheffield Hallam University, Faculty of Health and Wellbeing, Academy of Sport and

Physical Activity, Collegiate Hall, Sheffield, S10 2BP, United Kingdom

Corresponding author:

Linda van Gelder

l.v.gelder@shu.ac.uk

The Centre for Sports Engineering Research (CSER)

Faculty of Health and Wellbeing

Sheffield Hallam University

11 Broomgrove Road

S10 2LX Sheffield, United Kingdom

Word count:

Abstract: 222

Main text: 4425

Title:

The use of biofeedback for gait retraining: a mapping review

Abstract:

Background: Biofeedback seems to be a promising tool to improve gait outcomes for both healthy individuals and patient groups. However, due to differences in study designs and outcome measurements, it remains uncertain how different forms of feedback affect gait outcomes. Therefore, the aim of this study is to review primary biomechanical literature which has used biofeedback to alter gait-related outcomes in human participants.

Methods: Medline, Cinahl, Cochrane, SPORTDiscus and Pubmed were searched from inception to December 2017 using various keywords and the following MeSHterms: biofeedback, feedback, gait, walking and running. From the included studies, sixteen different study characteristics were extracted.

Findings: In this mapping review 173 studies were included. The most common feedback mode used was visual feedback (42%, n=73) and the majority fed-back kinematic parameters (36%, n=62). The design of the studies were poor: only 8% (n=13) of the studies had both a control group and a retention test; 69% (n=120) of the studies had neither. A retention test after 6 months was performed in 3% (n=5) of the studies, feedback was faded in 9% (n=15) and feedback was given in the field rather than the laboratory in 4% (n=8) of the studies.

Interpretation: Further work on biofeedback and gait should focus on the direct comparison between different modes of feedback or feedback parameters, along with better designed and field based studies.

Keywords:

Gait; movement retraining; biofeedback; real-time feedback

1. Introduction

Patient groups with lower-limb musculoskeletal and neurological conditions experience gait limitations (Baram, 2013; James, 1992; Richards et al., 2016; Tate and Milner, 2010), such as reduced walking speed and distance (Baram, 2013; James, 1992; Richards et al., 2016; Tate and Milner, 2010). These limitations can have a major impact on patients' lives, as their daily living activities and social interactions are often affected (Baram and Miller, 2006). Other examples of gait limitations include insufficient foot clearance for patients with multiple sclerosis (Bregman et al., 2010) and stroke patients (Balaban and Tok, 2014), leading to increased risk of trips and falls, a reduced push-off power for patients with multiple sclerosis (Bregman et al., 2010) and diabetes (Mueller et al., 1994) and increased knee flexion or excessive knee extension during walking for stroke patients (Balaban and Tok, 2014) and **individuals** with cerebral palsy (Rodda and Graham, 2001). Healthy individuals might also display gait patterns that predispose them to chronic overuse injuries. Tibial stress injuries (Agresta and Brown, 2015) and patellofemoral pain (Cheung and Davis, 2011) are both common running injuries for which altered landing mechanics have been identified as key risk factors (Noehren et al., 2012). Such overuse injuries can cause significant disruption **to** training, a reduction in physical fitness as well as personal frustration (Clansey et al., 2014).

Treatment options to reduce the risk of overuse injuries in athletes and improve gait limitations in patients, range from the use of orthotic devices to surgical procedures on nerves or muscles (National Institute for Health and Care Excellence, 2016a, 2016b, 2013; Yeung and Yeung, 2001). Gait retraining, a non-invasive technique which focusses on the rehabilitation of gait by either muscle strengthening, treadmill training, neurodevelopmental techniques or intensive mobility exercises (Eng and Fang Tang, 2007), is an additional treatment option. Understanding how gait retraining may be used to benefit different patient

groups or reduce the risk of overuse injuries is an important step in developing non-invasive treatment plans or prevention strategies to help improve individual outcomes.

Biofeedback makes use of electronic equipment to provide the user with additional biological information, beyond that which is naturally available to them (Agresta and Brown, 2015; James, 1992; Tate and Milner, 2010). Advances in technology have made biofeedback systems more affordable and more accessible to researchers; as a result there has been an increase in the literature in this area over recent years. Research suggests biofeedback to be a promising tool used to complement gait retraining (Stanton et al., 2011; Tate and Milner, 2010) and improve outcomes among several patient groups (Baram, 2013; James, 1992; Richards et al., 2016). For instance, stroke patients decreased the number of knee hyperextensions and increased gait speed when they received feedback on their joint kinematics (Stanton et al., 2011). Biofeedback has also been found to be effective at altering gait patterns in healthy subjects (Agresta and Brown, 2015; Richards et al., 2016) and reducing injury risk factors in runners (Agresta and Brown, 2015). Agresta and Brown (2015) found in their systematic review that runners demonstrated reduced kinetic risk factors associated with tibial stress fracture when receiving feedback on their peak tibial accelerations over the course of a run. Despite this, other studies included in the review of Tate and Milner (2010) have failed to find the use of biofeedback in gait retraining to be an effective tool in improving gait outcomes. These conflicting results might be due to differences in study designs and the populations examined (Stanton et al., 2011; Tate and Milner, 2010).

It is suggested that presenting the feedback in the field results in a more representative experimental design (Brunswik, 1956; Araújo et al., 2007). A more representative

1 experimental design provides a better representation of the behavioural setting, which could
2 lead to more beneficial and representative results (Araújo et al., 2007). With respect to the
3 mode of feedback, researchers have suggested that multisensory feedback is superior to
4 separate modes (visual, auditory, sensory) of feedback, not only due to encoding the most
5 information but also due to the reduction of cognitive load associated with the separate
6 systems due to distribution of information processing (Sigrist et al., 2013). With respect to
7 the feedback parameter, feedback on knowledge of results might be more beneficial than
8 feedback on knowledge of performance (Winstein, 1991). Further, studies have suggested
9 that gradually removing feedback over time -fading the feedback- reduces the chances of
10 participants becoming dependent on the feedback, facilitating improved learning (Agresta
11 and Brown, 2015; Richards et al., 2016). Moreover, long term follow-up retention tests after
12 gait retraining are important to assess learning (Agresta and Brown, 2015; Tate and Milner,
13 2010). Studies in the literature differ in the choice of feedback parameters and mode of
14 feedback given, as well as the length of any retention period, which makes it difficult to draw
15 firm conclusions about the effectiveness of, and optimal strategies for, gait retraining
16 interventions. Advances in technology have made biofeedback systems more affordable and
17 more accessible to researchers; as a result, there has been a surge in the literature in this area
18 over recent years. Therefore, a mapping review of the biofeedback for gait retraining
19 literature is required to get a broader understanding of the studies, characterise what has been
20 done, and to identify what areas need future research.

21
22 The aim of this study was to review primary biomechanical literature which has used
23 biofeedback to alter gait-related outcomes in human participants. Areas of interest included
24 the mode of feedback, which parameters were fed-back, the intervention design and the
25 length of any retention period. We intend that this rigorous approach to evaluating the trends

in the area will help to inform future research in these key areas, to help provide clarity for the use of biofeedback for gait retraining applications.

2. Methods

2.1 Research design

This study used a mapping review approach; mapping reviews give an overview of the existing published research and can be used to obtain a better insight into the literature within a particular area (Booth et al., 2016). The results can be used to identify gaps in the literature and inform more specific future reviews and/or primary research studies. A mapping review searches the literature in a systematic way, but does not exclude articles based on quality. In the current mapping review the focus was on the methods used rather than the outcome.

2.2 Data sources and search strategy

The following databases were systematically searched from inception to December 2017: Medline (via EBSCOhost Research Databases), Cinahl (via EBSCOhost Research Databases), Cochrane, SPORTDiscus (via EBSCOhost Research Databases) and Pubmed. Searches used the following combination of MeSH terms: (biofeedback (psychology) OR feedback (sensory)) AND (gait OR walking OR running). The same terms were searched separately in: Title, Abstract and Subject/Keywords. An exception was the term feedback which was not searched in the different fields as the term is too broad and led to an unmanageable volume of results. Instead, a selection of terms was combined to make the search more specific to the area of interest: augmented feedback, real-time feedback, sensory feedback, proprioceptive feedback, vibrotactile feedback, tactile feedback, visual feedback, virtual feedback, auditory feedback and audio feedback. There were exceptions for the databases: Cinahl and SPORTDiscus, which did not have a separate MeSH term for feedback

(sensory), for these databases the other MeSH terms were searched together with the separate search terms. Since there was no separate field for Keywords/Subject in Pubmed, all fields were searched in this database. Furthermore reference lists were checked from all relevant reviews that were found and additional articles were identified.

2.3 Study selection

The primary researcher (LvG) selected articles based on the relevance of the title and abstract using the following inclusion criteria: (1) feedback was given on biological information beyond what was naturally available to the participants; (2) feedback was given on one or more gait related parameters (corresponding to the categories of 'Feedback parameter' in Table 1); (3) at least one of the tasks performed in the research was gait (4) the study aimed to modify one or more gait related parameters as opposed to, for example, testing the validity of a system; (5) feedback was given in real-time; (6) measurements were performed using technology as opposed to verbal feedback; (7) treatment did not involve a combination of biofeedback and another treatment; (8) the article was written in English and (9) the article gave sufficient information on all the items listed in Table 1. The full texts of all articles that were deemed potentially relevant were then checked by the primary researcher using the same inclusion criteria.

2.4 Data extraction of included articles

The primary researcher extracted the information of interest (Table 1) from all articles that met the inclusion criteria. When an article reported a study that covered more than one category, each category was considered separately. This could occur when more than one participant group was tested, for example healthy participants and participants who experienced a stroke, when more than one feedback mode was tested, for example one group

got auditory feedback and one group got visual feedback or when different parameters were fed-back, for example one group got feedback on knee angle while another group got feedback on knee moment. A second researcher (AB) reviewed a random sample of 10% of the articles at the start of the process to check the reliability of data extraction. Any disagreements between the researches were discussed and a consensus was sought with a third researcher (BH). This informed the final data extraction form which was used for all articles.

2.5 Study design categorisation

The final set of articles were assigned to four categories based on their research design: (A) the study had an experimental and a control group of at least ten participants per group and a retention test; (B) the study had an experimental and a control group of at least ten participants per group, but no retention test; (C) the study had no control group or a control group with less than ten persons per group and a retention test and (D) the study had no control group or a control group with less than ten persons per group and no retention test. A control group was defined as a group who received no intervention or an alternative (non-biofeedback) intervention at the same time as the experimental group received biofeedback. Ten participants per group was used as a cut off since this was recommended by Whitehead et al. (2016) for trials with a large effect size (0.8) with 90% power and two-sided 5% significance. A retention test was defined as a **test after one day or longer** during which participants had to walk or run without biofeedback.

Topics	Categories
Authors	Healthy, runners, stroke/hemiplegia, Parkinson's, incomplete spinal cord injuries, cerebral palsy, multiple sclerosis, amputees, diabetics, knee injuries, other (included: ibromyalgia syndrome; uncompensated unilateral vestibular loss; bilateral peripheral vestibular loss/areflexia; different neurological gait disorders; out patients referred to a geriatric falls and balance clinic; inpatient rehabilitation program; asymptomatic participants; orthopaedic surgery; chronic ankle instability; hip arthroplasty with trochanteric osteotomy; idiopathic bilateral peripheral neuropath and Charcot-marie-tooth-disease; toe walking and Parkinson or stroke; spina bifida; lower extremity disabilities)
Journal	
Year of publication	
Number of participants	
Participant group	
Mode of feedback	Visual, auditory, sensory, visual-auditory, visual-sensory, auditory-sensory, multisensory which is a combination of visual, auditory and sensory feedback
Feedback parameter	Spatiotemporal (included: stride width and symmetry, step length, stride length and symmetry, stance time, swing time, temporal symmetry in stance), kinematic (included: ankle, knee, hip, pelvis and trunk joint angles, foot contact angle, shank angle, foot progression angle, toe-out in stance phase, knee distance, minimum toe clearance, peak tibial acceleration, anterior-posterior and medial-lateral position of the subject's trunk, trunk sway and angular velocity), kinetic (included: ground reaction force, average loading rate, torque, pressure of the heel, pressure of the foot, centre of pressure, centre of mass, weight bearing, knee medial tibiofemoral contact force, peak vertical force on the cane during gait and human-machine interaction forces), muscle activation , physiological (included: heart rate, ventilation, VO ₂ and lower extremity temperature), combination
Feedback system	Force sensors fixed on participants, force plates fixed in place, optical motion capture system, motion capture system and force plates fixed in place, inertial measurement unit, electromyography systems, other (included: video camera, green screen; two sensors who have to be close to each other; electrogoniometer; position transducer; ultrasound; electrode to measure brain waves; biofeedback unit stabilizer, P pressure of muscles; Lokomat system (exoskeleton); Cycle-ergometer; heart rate monitor; thermal feedback system; motion capture and accelerometers; force plates and inertial sensors; EMG, 3D kinematics and instrumented treadmill, infrared, SPLnFFT Noise Meter)
Feedback in the laboratory or in the field	Laboratory, field, combination
Number of sessions	1, 2-5, 6-10, 11-20, >20, continuously wearing the device
Frequency of training	1 session, daily, twice a day, once a week, 2 times a week, 2-3 a week, 3 times a week, 4 times a week, 5 times a week, continuously wearing the device, unknown
Fading of the feedback	Yes, no
Retention test and if so, after what time	None, < 1 week, ≥ 1 week, ≥ 4 weeks, ≥ 3 months, ≥ 6 months
Test with or without feedback	With, without
Feedback on gait or another task	Feedback on gait, feedback on gait and another task
Outcome	Beneficial, no difference between an experimental and a control group or between a pre- and post- test, negative, no inferential statistics

1 **Table 1.** The fields that were extracted and in the second column the categories that were

2 found for each field.

3

3. Results

3.1 Search results

1316 articles were identified in Medline, 392 in Cinahl, 333 in Cochrane, 303 in SPORTDiscus and 1769 in Pubmed (Fig 1). After removing duplicates a total of 2165 articles were checked for relevance based on the title and abstract and 1674 articles were excluded. The full text of the remaining 491 articles was checked against the inclusion criteria and 143 articles were identified as relevant to the review. Five additional articles from the reference lists of the reviews identified were also included. Details of all articles included in this review (n=148) can be found in the supplementary material. These articles included a total of 173 studies, since some articles reported more than one study.

3.2 Overview of study characteristics

3.2.1 Year of publication

There has been an increase in published studies over recent years (Fig 2), with most studies published in 2016 (n=26) and 2017 (n=20). When considering older studies from 1977 until 1994, participants only received auditory feedback or a combination of auditory and visual feedback. Sensory feedback was first reported in 1994 and multimodal feedback was not reported until 2010. The use of motion capture systems in combination with biofeedback for gait was first reported in 2010.

3.2.2 Participant groups

A total of 2479 participants, across the 173 studies, were included - with a mean of 15.5 (range: 1-240) participants per study. Groups included healthy participants, runners (healthy or injured) and participants with various gait disorders, numbers and percentages are depicted above the groups in the figure (Table 1, Fig 3).

3.2.3 Feedback mode

A range of feedback modes and combinations of modes were used within the included studies (Table 1, Fig 4). **The most common mode of feedback used was visual.**

3.2.4 Feedback parameter

A range of feedback parameters were used in the included studies (Table 1, Fig 5). **Kinematic parameters were most frequently fed-back.**

3.2.5 Feedback system

A variety of feedback systems (Table 1) were used to provide biofeedback to participants. Force sensors fixed to the participants feet or shoes were most frequently used (28%, n=49), followed by optical motion capture systems (15%, n=26), inertial measurement units (15%, n=25), motion capture in combination with force platforms (11%, n=19), force platforms alone (9%, n=16) and electromyography systems (9%, n=15). Other approaches were adopted in 13% (n=23) of the included studies.

3.2.6 Laboratory or field based studies

Ninety six percent (n=165) of the included studies were performed in a laboratory, 2% (n=4) in the field and the remaining 2% (n=4) used feedback given in both field and laboratory settings.

3.2.7 Training strategy and retention

More than half of all studies (53%, n=92) reported only one gait retraining session in which the participants received biofeedback. Three percent (n=5) of the studies reported 2-5

1 sessions, 20% (n=34) 6-10 session, 16% (n=27) 11-20 sessions while only 6% (n=11) gave
2 the participants more than 20 sessions of feedback. In two percent (n=4) of cases participants
3 were constantly wearing the device for the duration of the intervention. .
4

5 When studies included several sessions, most studies included 3 training sessions per week
6 (n=24, 14%), 11% (n=19) included two sessions a week and 6% (n=11) of the studies
7 reported up to 5 sessions a week. Three percent (n=5) of the studies included one training
8 session a week, 3% (n=5) included four sessions a week, 2% (n=3) of the studies had daily
9 training sessions, 1% of the studies included 2-3 training sessions a week (n=2) and 1% of the
10 studies included training sessions twice a day (n=2). In 2% (n=4) of the studies participants
11 wore the devices continuously in the field. Four percent (n=6) of the studies did not report the
12 frequency of the feedback sessions.
13

14 Only 9% (n=15) of the studies faded the feedback over the course of the gait retraining
15 intervention. In nine of these studies the task duration increased over time and the duration of
16 the feedback decreased. The other six articles did not increase task duration, but did
17 progressively decrease the feedback. Decreasing the feedback was done by giving alternating
18 blocks of feedback and blocks of no feedback. In 10% (n=18) of the studies feedback was
19 given on gait in combination with another task, such as a postural balance task.
20

21 Forty four percent (n=76) of the studies had no retention test, so the re-test was completed
22 while participants were still receiving biofeedback. Thirty-two percent (n=55) had a retention
23 test within a week of the intervention finishing, 8% (n=15) completed a retention test after
24 more than a week and within 4 weeks, 10% (n=17) after 4 weeks and within 3 months, 3%

(n=5) after 3 months and within 6 months, while only 3% (n=5) completed a retention test 6 months or more after the intervention finished.

3.3 Outcomes

Sixty eight percent (n=118) of the studies reported beneficial outcomes related to one or more gait parameters, 20% (n=34) reported no difference between the experimental and control groups and/or pre- and post- test outcomes and 12% (n=21) did not report inferential statistics. Negative effects of biofeedback on gait parameters were not reported in any studies.

3.4 Study design categories

Based on the study design categories outlined in the methods, only 8% (n=13) of all studies were in category A, 8% (n=14) in category B, 15% (n=26) in category C with the remaining studies (69%, n=120) categorized as group D. Since all studies in category A had an experimental and a control group of at least ten participants and a retention test, these studies were considered in further detail.

Research in category A used a range of participant groups (Table 2) with the majority of studies using visual feedback (n=5, S25, S328, S50, S105, S122) followed by a combination of visual and auditory (n=4, S24, S33, S96-1, S96-2), auditory (n=3, S61, S77, S101) feedback and one article using multisensory feedback (S94).

Most of these studies (S24, S33, S38, S94, S96-1, S96-2, S101, S122) provided feedback on kinematic parameters. Seven of the studies in this category (S24, S61, S77, S96-1, S96-2, S101, S122) reported 18 feedback sessions or more while 2 studies (S38, S105) used only a

1 single feedback session. Two studies (S24, S25) faded the feedback given and only one study
2 (S61) gave feedback in the field. Only 4 (S25, S33, S96-2, S101) of the 13 (31%) studies
3 reported beneficial effects of gait retraining on their selected outcome variable. In 6 (S50,
4 S38, S61, S94, S96-1, S122) of the studies a significant difference was reported between the
5 baseline and retention tests, but no significant difference was reported between the
6 experimental and the control groups. The remaining studies (S24, S105, S77) reported no
7 difference between baseline and retention tests or between groups.

Author	Year	Participant group	Number of participants	Mode of feedback	Feedback parameter	Feedback system	Training time	Fading	Retention	Study outcomes	Feedback on gait	Lab or field
Carpinella S24	2017	Parkinsons disease	Exp: 17 Con: 20	Visual and auditory	Combination of kinematic variables	Six inertial sensors	20 sessions, 45 minutes each, 3 times a week	Yes	1 month post training	No difference: No pre-post differences for walking speed. Significant differences were found for balance measurements.	Yes	Lab
Chan S25	2017	Novice runners	Exp: 166 con: 154	Visual	Vertical ground-reaction force signal	Instrumented treadmill	8 sessions, 4 times a week for 2 weeks	Yes	12 months post training	Beneficial: Both significant differences between baseline and retention test and between the experimental and control group.	Yes	Lab
Clansey S33	2014	Recreational rearfoot striking male runners	Exp: 12 Con: 10	Visual and auditory	Peak tibial axial accelerations	A tri-axial accelerometer	6 sessions, 2 times a week for 3 weeks	No	1 month post training	Beneficial: Both significant differences between baseline and retention test and between the experimental and control group.	Yes	Lab
Creaby S38	2016	Healthy male runners	Exp: 11 Con: 11	Visual	Peak tibial axial accelerations	A tri-axial accelerometer	1 session, 10 min	No	1 week post training	No difference: No between-group differences	Yes	Lab
Druzicki S50	2016	Stroke	Exp: 25 Con: 25	Visual	Step length	Instrumented treadmill	10 sessions, 5 times a week for 2 weeks	No	6 months post training	No difference: No between-group differences	Yes	Lab
Ginis S61	2016	Patients with Parkinson's disease	Exp: 22 Con: 18	Auditory	Cadence, stride length, symmetry and gait speed	CuPiD smartphone application, inertial measurement units	18 sessions, 3 times a week for six weeks	No	4 weeks post training	No difference: No between-group differences	Yes	Field

1 **Table 2.** Key data extracted from the studies in category A. References can be found in the supplement. Exp = experimental, Con = control.

Author	Year	Participant group	Number of participants	Mode of feedback	Feedback parameter	Feedback system	Training time	Fading	Retention	Study outcomes	Feedback on gait	Lab or field
Hurkmans S77	2012	Total hip arthroplasty with trochanteric osteotomy	Exp: 18 Con: 20	Auditory	Peak vertical force for each footstep	Insole pressure system, Pedalert system	Once per day during the entire hospital stay (6-8 weeks)	No	3 weeks post training	No difference: No between-group differences and no difference between the different retention tests	Yes	Lab
Lim S94	2016	Healthy older adults	Exp: 18 Con: 18	Multisensory	Trunk sway	SwayStarTM d gyroscopes	6 sessions, 3 times a week for 2 weeks	No	1 month post training	No difference: No between-group differences	Yes and balance tasks	Lab
Mandel S96	1990 (1)	Hemiparetic stroke patients	Exp: 13 Con: 11	Visual and auditory	Muscle activation of pretibial and calf muscle	Electromyography system	24 sessions, frequency unknown	No	3 months post training	No difference: No between-group differences	Yes and sitting, standing walking	Lab
Mandel S96	1990 (2)	Hemiparetic stroke patients	Exp: 13 Con: 11	Visual and auditory	Ankle position	Electrogoniometer	24 sessions, frequency unknown	No	3 months post training	Beneficial: Both significant differences between baseline and retention test and between the experimental and control group.	Yes and sitting, long sitting, standing and walking	Lab
Morris S101	1992	Patients with genu recurvatum following stroke	Exp: 13 Con: 13	Auditory	Peak amplitude of knee hyperextension	Electrogoniometer	20 session, 5 times a week for 4 weeks	No	4 weeks post training	Beneficial: Both significant differences between baseline and retention test and between the experimental and control group.	Yes	Lab
Oude Lansink S105	2017	Healthy participants	Exp: 13 Con: 11	Visual	Step width	Motion capture system	1 session, 1 min feedback	No	7-10 days post training	No difference, no effect of the treatment	Yes	Lab
Segal S122	2015	Patients with knee osteoarthritis	Exp: 19 Con: 19	Visual	Different kinematic trunk pelvic	Motion capture system	24 sessions, twice a week for 3 months	No	12 months post training	No difference: No between-group differences	Yes	Lab

1 **Table 2 continued.** Key data extracted from the studies in category A. References can be found in the supplement. **Exp = experimental, Con =**

2 **control.**

4. Discussion

The aim of this study was to review primary biomechanical literature which has used biofeedback to alter gait-related outcomes in human participants. A total of 173 relevant studies were identified. Visual feedback was the most commonly used mode and feedback on kinematic parameters was most commonly used. The vast majority of studies were performed in a laboratory and reported only one feedback session, did not fade the feedback given and had no retention test. Sixty-nine percent of all studies suggested some beneficial effects of biofeedback on gait outcomes with no significant negative effects reported, however this percentage of beneficial results was lower in studies that both included a control group and a retention test (Category A articles).

Visual feedback was given most frequently in the studies included in this mapping review. In a systematic review on injured and healthy runners, different modes of feedback were found to be effective in reducing variables related to ground reaction forces, but no mode of feedback was identified as being superior (Agresta and Brown, 2015). This is important since some modes of feedback such as auditory and sensory may be more practicable for use in field-based biofeedback systems. It has previously been suggested that multisensory is superior to separate modes of feedback, not only due to presenting the most information but also due to the reduction of cognitive load associated with the separate systems due to distribution of information processing (Sigrist et al., 2013). Some of the included studies in this mapping review directly compared different feedback modes. Hirokawa and Matsumura (1989) and Shin and Chung (2017) found the best gait-related outcomes when using combined visual and auditory feedback, compared to each mode separately. However, it should be noted that different modes of feedback were used for different parameters: visual feedback for step length and auditory feedback for step duration. A study comparing visual,

sensory and combined visual and sensory feedback on stride length in participants with incomplete spinal cord injury, found combined visual and sensory feedback to give significantly better results than the two modes presented separately (Yen et al., 2014). In this mapping review, multisensory feedback was only reported in 4% (n=6) of the studies. Future research on the effectiveness of different modes of feedback is therefore needed to help establish optimum feedback strategies for gait retraining applications within different populations. This suggestion supports previous research which has identified the need for research studies which directly compare different modes of feedback to further our knowledge in this area (Agresta and Brown, 2015; Sienko et al., 2017).

Kinematic variables were most frequently fed-back in the studies included in this mapping review. A previous systematic review on gait retraining found biofeedback of kinematic, kinetic and spatiotemporal parameters to show more promise than feedback on muscle activity, resulting in moderate to large short-term treatment effects in different patient groups (Tate and Milner, 2010). Feedback on muscle activity might be less effective since this mode of feedback focusses towards knowledge of performance. By moving away from knowledge of results and moving more towards knowledge of performance the learning response might be reduced (Winstein, 1991). Some studies included in this review support the suggestion that feedback on muscle activation results in smaller effects than feedback on other parameters. Franz et al. (2014) found that feedback on ground reaction forces (kinetic parameters) increased propulsive ground reaction forces and gastrocnemius muscle activity during push-off, while feedback on muscle activity only had no effect on the same gait related outcomes. In another study, feedback on muscle activity of the pretibial and calf muscles had no effect on walking speed, while feedback on ankle angle during heel-off and swing through (kinematic parameter) had a beneficial effect on the same gait related outcome (Mandel et al.,

1990). However, a direct comparison between kinetic and kinematic parameters has not been reported in gait related studies, therefore it remains uncertain which group of variables may offer the best outcomes. A direct comparison between the different groups of parameters is needed to provide more insight into which parameter might be most effective at improving gait related outcomes.

Only 4 of the 173 studies gave feedback in the field, with a further 4 studies giving a combination of laboratory and field based training. Even though two previous reviews concluded that field based systems should be considered (Richards et al., 2016; Shull et al., 2014), to date the vast majority of published research is confined to laboratory settings. Presenting feedback in the field may facilitate the trend for healthcare to move away from a clinical model to a self-care model supported by technology (McCullagh et al., 2010), and it would also improve the representative design of experiments (Araújo et al., 2007). However, presenting feedback in the field does have some practical implementation issues. For example, visual feedback could be shown on a screen in the laboratory, but this would not be easily possible in the field. Auditory and sensory feedbacks are therefore easier to facilitate in field based settings.

Future research should also focus on the design of feedback interventions. Over half of the included studies reported one feedback training session. Since beneficial outcomes could be related to the duration of the intervention (Adamovich et al., 2009; Agresta and Brown, 2015), both the duration and number of sessions required for effective retraining should be explored. These findings are supported by a review of Gordt et al. (2017) on the effects of feedback of wearable sensor data on balance, gait and functional performance in both healthy and patient populations. These authors concluded that future randomised controlled trials

1 should be designed with adequate intervention periods to enhance learning. In the current
2 mapping review, only fifteen of the included studies used a faded feedback approach within
3 their intervention. By gradually removing feedback over time, it is suggested that participants
4 do not become dependent on the feedback, facilitating improved learning (Winstein, 1991).
5 The majority of studies in this review had no retention test or a short term retention test
6 within a week of the intervention finishing. Establishing the long term retention of any gait
7 related changes represents a crucial step in prescribing gait retraining interventions as an
8 effective alternate to existing treatment options (Agresta and Brown, 2015; Gordt et al. 2017,
9 Stanton et al., 2017; Tate and Milner, 2010). Further, only thirteen studies combined having a
10 retention test with having a control group. Of those thirteen studies, eleven studies reported
11 beneficial effects of gait retraining when comparing baseline values to the retention values,
12 four studies found significant differences between experimental and control groups.
13 Therefore, the use of biofeedback shows promising results, since it has the same or a better
14 effect compared to existing interventions, without the need for a health practitioner, or
15 several trips to the clinic if field based feedback could be applied. However, at present there
16 is a lack of well-designed studies that have established the long term efficacy of biofeedback
17 for use in gait retraining interventions. Therefore, future work should focus on higher quality
18 study designs, with a special focus on assessing the long term effects of any interventions.

19
20 This review has some limitations that are noteworthy: we used a selection of terms combined
21 with feedback (as stated in the methods, section 2.2), since feedback is too broad as a term
22 and would therefore have led to too many results. By using a selection of terms instead of
23 feedback, there is a possibility that we missed some articles. However, we covered the area
24 which we were interested in by a wide selection of terms and we further searched the
25 reference list of reviews we found as well to make sure no articles were missed. Another

1 limitation is the risk of publication bias, which might inflate the number of beneficial effects
2 reported for the main outcome. Publication bias could mean that studies are less likely to be
3 published when they have not found beneficial results. By choosing a mapping review instead
4 of a systematic review we chose not to assess quality, assessing of the quality could have
5 reduced the publication bias. However, in the current review the focus was on assessing the
6 body of literature on the use of biofeedback to alter gait-related outcomes and the methods
7 used; for this a mapping review was the most appropriate approach.

8 9 **5. Conclusion**

10 There is a growing body of research on the use of biofeedback in gait retraining. This mapping
11 review has identified several areas within the current body of research that warrant further
12 work. Future research should focus on direct comparisons between groups of parameters and
13 feedback modes for specific gait retraining applications. Furthermore, researchers should
14 seek to produce high quality well designed studies that explore the fading of feedback, the
15 appropriate number of sessions as well as include a control group as assessing the long-term
16 benefits of any intervention. Finally, researchers should seek to develop and assess the
17 efficacy of field-based gait retraining systems using experimental designs more representative
18 of real life situations.

19 20 **Declarations of interest statement:**

21 Declarations of interest: none

22 23 **Acknowledgements:**

24 This research did not receive any specific grant from funding agencies in the public,
25 commercial, or not-for-profit sectors.

References

- Adamovich, S. V, Fluet, G.G., Tunik, E., Merians, A.S., 2009. Sensorimotor training in virtual reality: a review. *NeuroRehabilitation* 25, 29–44. <https://doi.org/10.3233/NRE-2009-0497>
- Agresta, C., Brown, A., 2015. Gait Retraining for Injured and Healthy Runners Using Augmented Feedback: A Systematic Literature Review. *J. Orthop. Sports Phys. Ther.* 45, 576–584. <https://doi.org/10.2519/jospt.2015.5823>
- Araújo, D., Davids, K., Passos, P., 2007. Ecological validity, representative design, and correspondence between experimental task constraints and behavioral setting: comment on Rogers, Kadar, and Costall (2005). *Ecol. Psychol.* 19, 69–78. <https://doi.org/10.1080/10407410709336951>
- Balaban, B., Tok, F., 2014. Gait disturbances in patients with stroke. *PM R* 6, 635–642. <https://doi.org/10.1016/j.pmrj.2013.12.017>
- Baram, Y., 2013. Virtual Sensory Feedback for Gait Improvement in Neurological Patients. *Front. Neurol.* 4, Article 138. <https://doi.org/10.3389/fneur.2013.00138>
- Baram, Y., Miller, A., 2006. Virtual reality cues for improvement of gait in patients with multiple sclerosis. *Neurology.* 66, 178–181. <https://doi.org/10.1212/01.wnl.0000194255.82542.6b>
- Begg, R.K., Tirosh, O., Said, C.M., Sparrow, W. a, Steinberg, N., Levinger, P., Galea, M.P., 2014. Gait training with real-time augmented toe-ground clearance information decreases tripping risk in older adults and a person with chronic stroke. *Front. Hum. Neurosci.* 8, Article 243. <https://doi.org/10.3389/fnhum.2014.00243>
- Booth, A., Sutton, A., Papaioannou, D., 2016. Systematic approaches to a successful literature review. *SAGE.*

- 1 Bregman, D.J.J., de Groot, V., van Diggele, P., Meulman, H., Houdijk, H., Harlaar, J., 2010.
2 Polypropylene ankle foot orthoses to overcome drop-foot gait in central neurological
3 patients: A mechanical and functional evaluation. *Prosthet. Orthot. Int.* 34, 293–304.
4 <https://doi.org/10.3109/03093646.2010.495969>
- 5 Brunswik, E., 1956. Perception and the representative design of psychological experiments,
6 2nd ed. University of California Press, Berkeley, CA.
- 7 Cheung, R.T.H., Davis, I.S., 2011. Landing Pattern Modification to Improve Patellofemoral
8 Pain in Runners: A Case Series. *J. Orthop. Sports Phys. Ther.* 41, 914–919.
9 <https://doi.org/10.2519/jospt.2011.3771>
- 10 Clansey, A.C., Hanlon, M., Wallace, E.S., Nevill, A., Lake, M.J., 2014. Influence of Tibial
11 shock feedback training on impact loading and running economy. *Med. Sci. Sports*
12 *Exerc.* 46, 973–981. <https://doi.org/10.1249/MSS.0000000000000182>
- 13 Crowell, H.P., Davis, I.S., 2011. Gait retraining to reduce lower extremity loading in runners.
14 *Clin. Biomech.* 26, 78–83. <https://doi.org/10.1016/j.clinbiomech.2010.09.003>
- 15 Eng, J.J., Fang Tang, P., 2007. Gait training strategies to optimize walking ability in people
16 with stroke: A synthesis of the evidence. *Expert Rev Neurother* 7, 1417–1436.
17 <https://doi.org/10.1586/14737175.7.10.1417>
- 18 Franz, J.R., Maletis, M., Kram, R., 2014. Real-time feedback enhances forward propulsion
19 during walking in old adults. *Clin. Biomech.* 29, 68–74.
20 <https://doi.org/10.1016/j.clinbiomech.2013.10.018>
- 21 Gordt, K., Gerhardy, T., Najafi, B., Schwenk, M., 2017. Effects of Wearable Sensor-Based
22 Balance and Gait Training on Balance, Gait, and Functional Performance in Healthy and
23 Patient Populations: A Systematic Review and Meta-Analysis of Randomized
24 Controlled Trials. *Gerontology* 64, 74–89. <https://doi.org/10.1159/000481454>
- 25 Hirokawa, S., Matsumura, K., 1989. Biofeedback gait training system for temporal and

distance factors. *Med. Biol. Eng. Comput.* 27, 8–13.

<https://doi.org/https://doi.org/10.1007/BF02442163>

James, R., 1992. Biofeedback Treatment for Cerebral Palsy in Children and Adolescents: A Review. *Pediatr. Exerc. Sci.* 4, 198–212.

<https://doi.org/https://doi.org/10.1123/pes.4.3.198>

Mandel, A.R., Nymark, J.R., Balmer, S.J., Grinnell, D.M., O’Riain, M.D., 1990. Electromyographic versus rhythmic positional biofeedback in computerized gait retraining with stroke patients. *Arch. Phys. Med. Rehabil.* 71, 649–654.

McCullagh, P.J., Nugent, C.D., Zheng, H., Burns, W.P., Davies, R.J., Black, N.D., Wright, P., Hawley, M.S., Eccleston, C., Mawson, S.J., Mountain, G.A., 2010. Promoting Behaviour Change in Long Term Conditions Using a Self-management Platform, in: *Designing Inclusive Interactions*. Springer, London, pp. 229–238.

Mueller, M.J., Minor, S.D., Sahrman, S.A., Schaaf, J.A., Strube, M.J., 1994. Differences in the gait characteristics of patients with diabetes and peripheral neuropathy compared with age-matched controls. *Phys. Ther.* 74, 299–313. [https://doi.org/10.1016/S0966-6362\(98\)00015-0](https://doi.org/10.1016/S0966-6362(98)00015-0)

National Institute for Health and Care Excellence, 2016a. Spasticity in children and young people overview [WWW Document]. URL <https://pathways.nice.org.uk/pathways/spasticity-in-children-and-young-people>

National Institute for Health and Care Excellence, 2016b. Surgery for children and young people with spasticity [WWW Document]. URL <https://pathways.nice.org.uk/pathways/spasticity-in-children-and-young-people#path=view%3A/pathways/spasticity-in-children-and-young-people/surgery-for-children-and-young-people-with-spasticity.xml&content=view-node%3Anodes-selective-dorsal-rhizotomy>

- 1 National Institute for Health and Care Excellence, 2013. Movement difficulties after a stroke
- 2 [WWW Document]. URL
- 3 [https://pathways.nice.org.uk/pathways/stroke#path=view%3A/pathways/stroke/moveme](https://pathways.nice.org.uk/pathways/stroke#path=view%3A/pathways/stroke/movement-difficulties-after-a-stroke.xml&content=view-node%3Anodes-physiotherapy)
- 4 [nt-difficulties-after-a-stroke.xml&content=view-node%3Anodes-physiotherapy](https://pathways.nice.org.uk/pathways/stroke#path=view%3A/pathways/stroke/movement-difficulties-after-a-stroke.xml&content=view-node%3Anodes-physiotherapy)
- 5 Noehren, B., Pohl, M.B., Sanchez, Z., Cunningham, T., Lattermann, C., 2012. Proximal and
- 6 distal kinematics in female runners with patellofemoral pain. *Clin. Biomech.* 27, 366–
- 7 371. <https://doi.org/10.1016/j.clinbiomech.2011.10.005>
- 8 Richards, R., Van Den Noort, J., Dekker, J., Harlaar, J., 2016. Effects of gait retraining with
- 9 real-time biofeedback in patients with knee osteoarthritis: Systematic review and meta-
- 10 analysis. *Osteoarthr. Cartil.* 24, S470. <https://doi.org/10.1016/j.joca.2016.01.858>
- 11 Rodda, J., Graham, H.K., 2001. Classification of gait patterns in spastic hemiplegia and
- 12 spastic diplegia: a basis for a management algorithm. *Eur. J. Neurol.* 8, 98–108.
- 13 <https://doi.org/10.1046/j.1468-1331.2001.00042.x>
- 14 Shin, J., Chung, Y., 2017. Influence of visual feedback and rhythmic auditory cue on walking
- 15 of chronic stroke patient induced by treadmill walking in real-time basis.
- 16 *NeuroRehabilitation* 41, 445–452. <https://doi.org/10.3233/NRE-162139>
- 17 Shull, P.B., Jirattigalachote, W., Hunt, M.A., Cutkosky, M.R., Delp, S.L., 2014. Quantified
- 18 self and human movement: A review on the clinical impact of wearable sensing and
- 19 feedback for gait analysis and intervention. *Gait Posture* 40, 11–19.
- 20 <https://doi.org/10.1016/j.gaitpost.2014.03.189>
- 21 Sienko, K.H., Whitney, S.L., Carender, W.J., Wall, C., 2017. The role of sensory
- 22 augmentation for people with vestibular deficits: Real-time balance aid and/or
- 23 rehabilitation device? *J. Vestib. Res. Equilib. Orientat.* 27, 63–76.
- 24 <https://doi.org/10.3233/VES-170606>
- 25 Sigrist, R., Rauter, G., Riener, R., Wolf, P., 2013. Augmented visual, auditory, haptic, and

multimodal feedback in motor learning: A review. *Psychon. Bull. Rev.* 20, 21–53.
<https://doi.org/10.3758/s13423-012-0333-8>

Stanton, R., Ada, L., Dean, C.M., Preston, E., 2017. Biofeedback improves performance in lower limb activities more than usual therapy in people following stroke: a systematic review. *J. Physiother.* 63, 11–16. <https://doi.org/10.1016/j.jphys.2016.11.006>

Stanton, R., Ada, L., Dean, C.M., Preston, E., 2011. Biofeedback improves activities of the lower limb after stroke: A systematic review. *J. Physiother.* 57, 145–155.
[https://doi.org/10.1016/S1836-9553\(11\)70035-2](https://doi.org/10.1016/S1836-9553(11)70035-2)


Tate, J.J., Milner, C.E., 2010. Real-Time Kinematic, Temporospacial, and Kinetic Biofeedback During Gait Retraining in Patients: A Systematic Review. *Phys. Ther.* 90, 1123–1134. <https://doi.org/10.2522/ptj.20080281>


Whitehead, A.L., Julious, S.A., Cooper, C.L., Campbell, M.J., 2016. Estimating the sample size for a pilot randomised trial to minimise the overall trial sample size for the external pilot and main trial for a continuous outcome variable. *Stat. Methods Med. Res.* 25, 1057–1073. <https://doi.org/10.1177/0962280215588241>


Winstein, C.J., 1991. Knowledge of results and motor learning - Implications for physical therapy. *Phys. Ther.* 71, 140–149. <https://doi.org/10.1093/ptj/71.2.140>


Yen, S.C., Landry, J.M., Wu, M., 2014. Augmented multisensory feedback enhances locomotor adaptation in humans with incomplete spinal cord injury. *Hum. Mov. Sci.* 35, 80–93. <https://doi.org/10.1016/j.humov.2014.03.006>


Yeung, E.W., Yeung, S.S., 2001. A systematic review of interventions to prevent lower limb soft tissue running injuries. *Br. J. Sports Med.* 35, 383–389.
<https://doi.org/10.1136/bjism.35.6.383>











Declarations of interest statement:

Declarations of interest: none

Figure legends

Fig 1. Flow diagram of search strategy

Fig 2. Number of studies published each year

Fig 3. The number of studies published for each participant group

Fig 4. The number of studies published for each mode of feedback

Fig 5. The numbers of studies published for each parameter which was fed-back

References

- [S1] M.R. Afzal, M.-K. Oh, C.-H. Lee, Y.S. Park, J. Yoon, M.R. Afzal, M.-K. Oh, C.-H. Lee, Y.S. Park, J. Yoon, A Portable Gait Asymmetry Rehabilitation System for Individuals with Stroke Using a Vibrotactile Feedback., *Biomed Res. Int.* 2015 (2015) 1-16. doi:10.1155/2015/375638.
- [S2] E. Aiello, D.H. Gates, B.L. Patritti, K.D. Cairns, M. Meister, E. a Clancy, P. Bonato, Visual EMG Biofeedback to Improve Ankle Function in Hemiparetic Gait., *Conf. Proc. IEEE Eng. Med. Biol. Soc.* 7 (2005) 7703–7706. doi:10.1109/IEMBS.2005.1616297.
- [S3] J.E. Aikens, Thermal Biofeedback For Claudication In Diabetes: A Literature Review and Case Study, *Altern Med Rev.* 4 (1999) 104–110.
- [S4] J.H.J. Allum, M.G. Carpenter, B.C. Horslen, J.R. Davis, F. Honegger, K.-S. Kok-Sing Tang, P. Kessler, Improving impaired balance function: Real-time versus carry-over effects of prosthetic feedback, in: 2011 Annu. Int. Conf. IEEE Eng. Med. Biol. Soc., IEEE, 2011: pp. 1314–1318. doi:10.1109/IEMBS.2011.6090309.
- [S5] E. Anson, R. Rosenberg, P. Agada, T. Kiemel, J. Jeka, Does visual feedback during walking result in similar improvements in trunk control for young and older healthy adults?, *J. Neuroeng. Rehabil.* 10 (2013) 110. doi:10.1186/1743-0003-10-110.
- [S6] A.S. Aruin, T.A. Hanke, A. Sharma, Base of support feedback in gait rehabilitation., *Int. J. Rehabil. Res.* 26 (2003) 309–312. doi:10.1097/01.mrr.0000102059.48781.a8.
- [S7] A.S. Aruin, A. Sharma, R. Larkins, G. Chaudhuri, Knee position feedback: its effect on management of pelvic instability in a stroke patient, *Disabil. Rehabil.* 22 (2000) 690–692. doi:10.1080/096382800445498.
- [S8] J. Bae, K. Kong, N. Byl, M. Tomizuka, A Mobile Gait Monitoring System for Abnormal Gait Diagnosis and Rehabilitation: A Pilot Study for Parkinson Disease Patients, *J. Biomech. Eng.* 133 (2011) 41005. doi:10.1115/1.4003525.
- [S9] M. Baggaley, R.W. Willy, S.A. Meardon, Primary and secondary effects of real-time feedback to reduce vertical loading rate during running, *Scand. J. Med. Sci. Sports.* (2016) 1–7. doi:10.1111/sms.12670.
- [S10] R. Banz, M. Bolliger, G. Colombo, V. Dietz, L. Lünenburger, Computerized visual feedback: an adjunct to robotic-assisted gait training., *Phys. Ther.* 88 (2008) 1135–1145. doi:10.2522/ptj.20070203.
- [S11] J.A. Barrios, K.M. Crossley, I.S. Davis, Gait retraining to reduce the knee adduction moment through real-time visual feedback of dynamic knee alignment, *J. Biomech.* 43 (2010) 2208–2213. doi:10.1016/j.jbiomech.2010.03.040.
- [S12] N. Basaglia, N. Mazzini, P. Boldrini, P. Bacciglieri, E. Contenti, G. Ferraresi, Biofeedback treatment of genu-recurvatum using an electrogoniometric device with an acoustic signal. One-year follow-up, *Scand. J. Rehabil. Med.* 21 (1989) 125–130.
- [S13] M. Batavia, J.G. Gianutsos, A. Vaccaro, J.T. Gold, A do-it-yourself membrane-activated auditory feedback device for weight bearing and gait training: A case report, *Arch. Phys. Med. Rehabil.* 82 (2001) 541–545. doi:10.1053/apmr.2001.21931.
- [S14] R.K. Begg, O. Tirosh, C.M. Said, W. a Sparrow, N. Steinberg, P. Levinger, M.P. Galea, Gait training with real-time augmented toe-ground clearance information decreases tripping risk in older adults and a person with chronic stroke., *Front. Hum. Neurosci.* 8 (2014) Article 243. doi:10.3389/fnhum.2014.00243.
- [S15] J. Berengueres, M. Fritschi, R. McClanahan, A smart pressure-sensitive insole that reminds you to walk correctly: an orthotic-less treatment for over pronation, *Eng. Med. Biol. Soc. (EMBC), 2014 36th Annu. Int. Conf. IEEE.* (2014) 2488–2491.

doi:10.1109/EMBC.2014.6944127.

- [S16] S.A. Binder, C.B. Moll, S.L. Wolf, Evaluation of electromyographic biofeedback as an adjunct to therapeutic exercise in treating the lower extremities of hemiplegic patients., *Phys. Ther.* 61 (1981) 886–893.
- [S17] J.E. Bolek, A preliminary study of modification of gait in real-time using surface electromyography, *Appl. Psychophysiol. Biofeedback.* 28 (2003) 129–138. doi:10.1023/A:1023810608949.
- [S18] L. Bradley, B. Hart, S. Mandana, K. Flowers, M. Riches, P. Sanderson, Electromyographic biofeedback for gait training after stroke, *Clin. Rehabil.* 12 (1998) 11–22. doi:10.1191/026921598677671932.
- [S19] A. Brasileiro, G. Gama, L. Trigueiro, T. Ribeiro, E. Silva, É. Galvão, A. Lindquist, Influence of visual and auditory biofeedback on partial body weight support treadmill training of individuals with chronic hemiparesis: a randomized controlled clinical trial., *Eur. J. Phys. Rehabil. Med.* 51 (2015) 49–58.
- [S20] K. Brüttsch, A. Koenig, L. Zimmerli, S. Mérillat-Koeneke, R. Riener, L. Jäncke, H.J.A. van Hedel, A. Meyer-Heim, Virtual reality for enhancement of robot-assisted gait training in children with central gait disorders., *J. Rehabil. Med.* 43 (2011) 493–9. doi:10.2340/16501977-0802.
- [S21] M.A. Busa, J. Lim, R.E.A. van Emmerik, J. Hamill, Head and Tibial Acceleration as a Function of Stride Frequency and Visual Feedback during Running., *PLoS One.* 11 (2016) e0157297. doi:10.1371/journal.pone.0157297.
- [S22] N. Byl, W. Zhang, S. Coo, M. Tomizuka, Clinical impact of gait training enhanced with visual kinematic biofeedback: Patients with Parkinson’s disease and patients stable post stroke, *Neuropsychologia.* 79 (2015) 332–343. doi:10.1016/j.neuropsychologia.2015.04.020.
- [S23] G.G. Caird, S. J., McKenzie, A. D., & Sleivert, Biofeedback and relaxation techniques improve running economy in sub-elite long distance runners., *Med. Sci. Sports Exerc.* 31 (1999) 717–722.
- [S24] I. Carpinella, D. Cattaneo, G. Bonora, T. Bowman, L. Martina, A. Montesano, M. Ferrarin, Wearable Sensor-Based Biofeedback Training for Balance and Gait in Parkinson Disease: A Pilot Randomized Controlled Trial, *Arch. Phys. Med. Rehabil.* 98 (2017) 622–630. doi:http://dx.doi.org/10.1016/j.apmr.2016.11.003
- [S25] Z.Y.S. Chan, J.H. Zhang, I.P.H. Au, W.W. An, G.L.K. Shum, G.Y.F. Ng, R.T.H. Cheung, Gait Retraining for the Reduction of Injury Occurrence in Novice Distance Runners: 1-Year Follow-up of a Randomized Controlled Trial, *Am. J. Sports Med.* (2017) 36354651773627. doi:10.1177/0363546517736277.
- [S26] T.L. Chen, W.W. An, Z.Y.S. Chan, I.P.H. Au, Z.H. Zhang, R.T.H. Cheung, Immediate effects of modified landing pattern on a probabilistic tibial stress fracture model in runners, *Clin. Biomech.* 33 (2016) 49–54. doi:http://dx.doi.org/10.1016/j.clinbiomech.2016.02.013.
- [S27] D.K.Y. Chen, M. Haller, T.F. Besier, Wearable lower limb haptic feedback device for retraining Foot Progression Angle and Step Width, *Gait Posture.* 55 (2017) 177–183. doi:10.1016/j.gaitpost.2017.04.028.
- [S28] Y.-L. Chen, Y.C. Li, T.-S.S. Kuo, J.-S.S. Lai, The development of a closed-loop controlled functional electrical stimulation (FES) in gait training., *J. Med. Eng. Technol.* 25 (2001) 41–48. doi:10.1080/03091900110043612.
- [S29] R.T.H. Cheung, W.W. An, I.P.H. Au, J.H. Zhang, Z.Y.S. Chan, A.J. MacPhail, Control of impact loading during distracted running before and after gait retraining in runners, *J. Sports Sci.* (2017) 1–5. doi:10.1080/02640414.2017.1398886.

- [S30] R.T.H.. Cheung, I.S. Davis, Landing Pattern Modification to Improve Patellofemoral Pain in Runners: A Case Series, *J. Orthop. Sports Phys. Ther.* 41 (2011) 914–919. doi:10.2519/jospt.2011.3771.
- [S31] Y.H. Choi, J.D. Kim, J.H. Lee, Y.J. Cha, Walking and balance ability gain from two types of gait intervention in adult patients with chronic hemiplegic stroke: A pilot study, *Assist. Technol.* (2017). doi:10.1080/10400435.2017.138761
- [S32] I. Cikajlo, Z. Matjacic, T. Bajd, R. Futami, Sensory Supported FES Control in Gait Training of Incomplete Spinal Cord Injury Persons, *Artif. Organs.* 29 (2005) 459–461. doi:10.1111/j.1525-1594.2005.29077.x.
- [S33] A.C. Clansey, M. Hanlon, E.S. Wallace, A. Nevill, M.J. Lake, Influence of Tibial shock feedback training on impact loading and running economy, 2014. doi:10.1249/MSS.0000000000000182.
- [S34] G.R. Colborne, S.J. Olney, M.P. Griffin, Feedback of ankle joint angle and soleus electromyography in the rehabilitation of hemiplegic gait, *Arch. Phys. Med. Rehabil.* 74 (1993) 1100–1106. doi:10.1016/0003-9993(93)90069-M.
- [S35] G.R. Colborne, F. V Wright, S. Naumann, Feedback of triceps surae EMG in gait of children with cerebral palsy: a controlled study, *Arch. Phys. Med. Rehabil.* 75 (1994) 40–45.
- [S36] E.E. Conrad, Libby; Bleck, Augmented Auditory Feedback in the Treatment of Equinus Gait in Children, *Dev. Med. Child Neurol.* 22 (1980) 713–718. doi:10.1111/j.1469-8749.1980.tb03737.x.
- [S37] S. Crea, B.B. Edin, K. Knaepen, R. Meeusen, N. Vitiello, Time-Discrete Vibrotactile Feedback Contributes to Improved Gait Symmetry in Patients With Lower Limb Amputations: Case Series, *Phys. Ther.* 97 (2017) 198–207. doi:10.2522/ptj.20150441.
- [S38] M.W. Creaby, M.M. Franettovich Smith, Retraining running gait to reduce tibial loads with clinician or accelerometry guided feedback, *J. Sci. Med. Sport.* 19 (2016) 288–292. doi:10.1016/j.jsams.2015.05.003.
- [S39] H.P. Crowell, I.S. Davis, Gait retraining to reduce lower extremity loading in runners, *Clin. Biomech.* 26 (2011) 78–83. doi:10.1016/j.clinbiomech.2010.09.003.
- [S40] H.P. Crowell, C.E. Milner, J. Hamill, I.S. Davis, Reducing impact loading during running with the use of real-time visual feedback., *J. Orthop. Sports Phys. Ther.* 40 (2010) 206–213. doi:10.2519/jospt.2010.3166.
- [S41] B.J. Darter, J.M. Wilken, Gait training with virtual reality-based real-time feedback: improving gait performance following transfemoral amputation., *Phys. Ther.* 91 (2011) 1385–1394. doi:10.2522/ptj.20100360.
- [S42] J.R. Davis, M.G. Carpenter, R. Tschanz, S. Meyes, D. Debrunner, J. Burger, J.H.J. Allum, Trunk sway reductions in young and older adults using multi-modal biofeedback, *Gait Posture.* 31 (2010) 465–472. doi:10.1016/j.gaitpost.2010.02.002.
- [S43] B.L. Davis, M. Ortolano, K. Richards, J. Redhed, J. Kuznicki, V. Sahgal, Realtime visual feedback diminishes energy consumption of amputee subjects during treadmill locomotion, *J. Prosthetics Orthot.* 16 (2004) 49–54.
- [S44] D. De León Rodriguez, L. Allet, A. Golay, J. Philippe, J.-P. Assal, C.-A. Hauert, Z. Pataky, Biofeedback can reduce foot pressure to a safe level and without causing new at-risk zones in patients with diabetes and peripheral neuropathy., *Diabetes. Metab. Res. Rev.* 29 (2013) 139–144. doi:10.1002/dmrr.2366.
- [S45] A. Descatoire, A. Thévenon, P. Moretto, Baropodometric information return device for foot unloading, *Med. Eng. Phys.* 31 (2009) 607–613. doi:10.1016/j.medengphy.2008.12.002.
- [S46] J.B. Dingwell, B.L. Davis, D.M. Frazier, Use of an instrumented treadmill for real-

time gait symmetry evaluation and feedback in normal and trans-tibial amputee subjects., *Prosthet. Orthot. Int.* 20 (1996) 101–110.
doi:10.3109/03093649609164426.

- [S47] L. Donovan, M.A. Feger, J.M. Hart, S. Saliba, J. Park, J. Hertel, Effects of an auditory biofeedback device on plantar pressure in patients with chronic ankle instability., *Gait Posture*. 44 (2016) 29–36. doi:10.1016/j.gaitpost.2015.10.013.
- [S48] A. V. Dowling, D.S. Fisher, T.P. Andriacchi, Gait Modification via Verbal Instruction and an Active Feedback System to Reduce Peak Knee Adduction Moment, *J. Biomech. Eng.* 132 (2010) 71007. doi:10.1115/1.4001584.
- [S49] M. Druzbicki, A. Guzik, G. Przysada, A. Kwolek, A. Brzozowska-Magoń, Efficacy of gait training using a treadmill with and without visual biofeedback in patients after stroke: A randomized study, *J. Rehabil. Med.* 47 (2015) 419–425.
doi:10.2340/16501977-1949.
- [S50] M. Druzbicki, A. Guzik, G. Przysada, A. Kwolek, A. Brzozowska-Magoń, M. Sobolewski, Changes in Gait Symmetry After Training on a Treadmill with Biofeedback in Chronic Stroke Patients: A 6-Month Follow-Up From a Randomized Controlled Trial, *Med. Sci. Monit.* 22 (2016) 4859–4868.
doi:10.12659/MSM.898420.
- [S51] M. Druzbicki, A. Kwolek, A. Depa, G. Przysada, The use of a treadmill with biofeedback function in assessment of relearning walking skills in post-stroke hemiplegic patients – a preliminary report, *Neurol. Neurochir. Pol.* 44 (2010) 567–573. doi:10.1016/S0028-3843(14)60154-7.
- [S52] M.S. El-Tamawy, M.H. Darwish, M.E. Khallaf, Effects of augmented proprioceptive cues on the parameters of gait of individuals with Parkinson’s disease., *Ann. Indian Acad. Neurol.* 15 (2012) 267–72. doi:10.4103/0972-2327.104334.
- [S53] M. Eriksson, K. a Halvorsen, L. Gullstrand, Immediate effect of visual and auditory feedback to control the running mechanics of well-trained athletes., *J. Sports Sci.* 29 (2011) 253–262. doi:10.1080/02640414.2010.523088.
- [S54] E.R. Esposito, H.S. Choi, B.J. Darter, J.M. Wilken, Can real-time visual feedback during gait retraining reduce metabolic demand for individuals with transtibial amputation?, *PLoS One*. 12 (2017) 1–14. doi:10.1371/journal.pone.0171786.
- [S55] J. Feasel, M.C. Whitton, S. Member, L. Kassler, S. Member, F.P.B. Jr, L. Fellow, M.D. Lewek, The Integrated Virtual Environment Rehabilitation Treadmill System, *IEEE Trans. NEURAL Syst. Rehabil. Eng.* 19 (2011) 290–297.
- [S56] V.G. Femery, P.G. Moretto, J.-M.G. Hespel, A. Thévenon, G. Lensel, A real-time plantar pressure feedback device for foot unloading, *Arch. Phys. Med. Rehabil.* 85 (2004) 1724–1728. doi:10.1016/j.apmr.2003.11.031.
- [S57] C. Ferrigno, I.S. Stoller, N. Shakoar, L.E. Thorp, M.A. Wimmer, The Feasibility of Using Augmented Auditory Feedback From a Pressure Detecting Insole to Reduce the Knee Adduction Moment: A Proof of Concept Study, *J. Biomech. Eng.* 138 (2016) 21014. doi:10.1115/1.4032123.
- [S58] J.R. Franz, M. Maletis, R. Kram, Real-time feedback enhances forward propulsion during walking in old adults, *Clin. Biomech.* 29 (2014) 68–74.
doi:10.1016/j.clinbiomech.2013.10.018.
- [S59] M.C. Fu, L. DeLuke, R. a Buerba, R.E. Fan, Y.J. Zheng, M.P. Leslie, M.R. Baumgaertner, J.N. Grauer, Haptic biofeedback for improving compliance with lower-extremity partial weight bearing., *Orthopedics*. 37 (2014) e993–e998.
doi:10.3928/01477447-20141023-56.
- [S60] K. Ghoseiri, B. Forogh, M.A. Sanjari, A. Bavi, The effect of a vibratory lumbar orthosis on walking velocity in patients with Parkinson’s disease., *Prosthet. Orthot.*

- Int. 33 (2009) 82–88. doi:10.1080/03093640802647094.
- [S61] P. Ginis, A. Nieuwboer, M. Dorfman, A. Ferrari, E. Gazit, C.G. Canning, L. Rocchi, L. Chiari, J.M. Hausdorff, A. Mirelman, Feasibility and effects of home-based smartphone-delivered automated feedback training for gait in people with Parkinson's disease: A pilot randomized controlled trial, *Parkinsonism Relat. Disord.* 22 (2016) 28–34. doi:10.1016/j.parkreldis.2015.11.004.
 - [S62] P.R. Golyski, E.M. Bell, E.M. Husson, E.J. Wolf, B.D. Hendershot, Modulation of Vertical Ground Reaction Impulse with Real-Time Biofeedback: A Feasibility Study, *J. Appl. Biomech.* 32 (2017) 1–23. doi:10.1123/jab.2017-0004.
 - [S63] K.E. Gordon, D.P. Ferris, A.D. Kuo, Metabolic and Mechanical Energy Costs of Reducing Vertical Center of Mass Movement During Gait, *Arch. Phys. Med. Rehabil.* 90 (2009) 136–144. doi:10.1016/j.apmr.2008.07.014.
 - [S64] M. Gray, E. Sweeney, M. Creaby, M. Smith, Gait retraining using visual and verbal feedback in runners, *30Th Annu. Conf. Biomech. Sport.* (2012) 262–263.
 - [S65] L. Hak, H. Houdijk, P.J. Beek, J.H. Van Dieën, Steps to take to enhance gait stability: The effect of stride frequency, stride length, and walking speed on local dynamic stability and margins of stability, *PLoS One.* 8 (2013). doi:10.1371/journal.pone.0082842.
 - [S66] L. Hak, H. Houdijk, P. Van Der Wurff, M.R. Prins, P.J. Beek, J.H. Van Dieën, Stride frequency and length adjustment in post-stroke individuals: Influence on the margins of stability, *J. Rehabil. Med.* 47 (2015) 126–132. doi:10.2340/16501977-1903.
 - [S67] D. Hamacher, D. Bertram, C. Fölsch, L. Schega, Evaluation of a visual feedback system in gait retraining: A pilot study, *Gait Posture.* 36 (2012) 182–186. doi:10.1016/j.gaitpost.2012.02.012.
 - [S68] N. Hegde, G.D. Fulk, E.S. Sazonov, Development of the RT-GAIT, a Real-Time feedback device to improve Gait of individuals with stroke, *Proc. Annu. Int. Conf. IEEE Eng. Med. Biol. Soc. EMBS.* 2015–Novem (2015) 5724–5727. doi:10.1109/EMBC.2015.7319692.
 - [S69] J. Hegeman, F. Honegger, M. Kupper, J.H.J. Allum, The balance control of bilateral peripheral vestibular loss subjects and its improvement with auditory prosthetic feedback., *J. Vestib. Res.* 15 (2005) 109–117.
 - [S70] E. Hershko, C. Tauber, E. Carmeli, Biofeedback versus physiotherapy in patients with partial weight-bearing, *Am. J. Orthop.* 37 (2008) E92–E96.
 - [S71] S. Hirokawa, K. Matsumura, Biofeedback gait training system for temporal and distance factors, *Med. Biol. Eng. Comput.* 27 (1989) 8–13.
 - [S72] R.E. Hogue, S. McCandless, Genu Recurvatum: Auditory Biofeedback Treatment for Adult Patients with Stroke or Head Injuries, *Arch. Phys. Med. Rehabil.* 64 (1983) 368–370.
 - [S73] S. Huang, J.P. Wensman, D.P. Ferris, Locomotor Adaptation by Transtibial Amputees Walking with an Experimental Powered Prosthesis Under Continuous Myoelectric Control, *IEEE Trans. Neural Syst. Rehabil. Eng.* 24 (2016) 573–581. doi:10.1109/TNSRE.2015.2441061.
 - [S74] M.A. Hunt, M. Simic, R.S. Hinman, K.L. Bennell, T. V. Wrigley, Feasibility of a gait retraining strategy for reducing knee joint loading: Increased trunk lean guided by real-time biofeedback, *J. Biomech.* 44 (2011) 943–947. doi:10.1016/j.jbiomech.2010.11.027.
 - [S75] M.A. Hunt, J. Takacs, K. Hart, E. Massong, K. Fuchko, J. Biegler, Comparison of mirror, raw video, and real-time visual biofeedback for training toe-out gait in individuals with knee osteoarthritis, *Arch. Phys. Med. Rehabil.* 95 (2014) 1912–

1917. doi:10.1016/j.apmr.2014.05.016.
- [S76] L. Hunter, Q.A. Louw, S.-M. van Niekerk, Effect of running retraining on pain, function, and lower-extremity biomechanics in a female runner with iliotibial band syndrome, *J. Sport Rehabil.* 23 (2014) 145–157. doi:10.1123/jsr.2013-0024.
 - [S77] H.L. Hurkmans, J.B. Bussmann, E. Benda, J.A. Verhaar, H.J. Stam, Effectiveness of audio feedback for partial weight-bearing in and outside the hospital: A randomized controlled trial, *Arch. Phys. Med. Rehabil.* 93 (2012) 565–570. doi:10.1016/j.apmr.2011.11.019.
 - [S78] J.W. Hustedt, D.J. Blizzard, M.R. Baumgaertner, M.P. Leslie, J.N. Grauer, Effect of age on partial weight-bearing training., *Orthopedics.* 35 (2012) e1061–e1067. doi:10.3928/01477447-20120621-23.
 - [S79] D. Intiso, V. Santilli, M.G. Grasso, R. Rossi, I. Caruso, Rehabilitation of walking with electromyographic biofeedback in foot-drop after stroke., *Stroke.* 25 (1994) 1189–1192. doi:10.1161/01.STR.25.6.1189.
 - [S80] E. Isakov, Gait rehabilitation: a new biofeedback device for monitoring and enhancing weight-bearing over the affected lower limb, *Eura Medicophys.* 43 (2007) 21–26.
 - [S81] B. Jackson, K.E. Gordon, A.H. Chang, Immediate and short-term effects of real-time knee adduction moment feedback on the peak and cumulative knee load during walking, *J. Orthop. Res.* (2017) 1–8. doi:10.1002/jor.23659.
 - [S82] L.J.F. Janssen, L.L. Verhoeff, C.G.C. Horlings, J.H.J. Allum, Directional effects of biofeedback on trunk sway during stance tasks in healthy young adults, *Gait Posture.* 29 (2009) 575–581. doi:10.1016/j.gaitpost.2008.12.009.
 - [S83] J. Jellish, J.J. Abbas, T.M. Ingalls, P. Mahant, J. Samanta, M.C. Ospina, N. Krishnamurthi, A System for Real-Time Feedback to Improve Gait and Posture in Parkinson ' s Disease, *IEEE J. Biomed. Heal. Informatics.* 19 (2015) 1809–1819. doi:10.1109/JBHI.2015.2472560.
 - [S84] J. Jonsdottir, D. Cattaneo, M. Recalcati, A. Regola, M. Rabuffetti, M. Ferrarin, A. Casiraghi, Task-Oriented Biofeedback to Improve Gait in Individuals With Chronic Stroke: Motor Learning Approach, *Neurorehabil. Neural Repair.* 24 (2010) 478–485. doi:10.1177/1545968309355986.
 - [S85] J. Jonsdottir, D. Cattaneo, A. Regola, A. Crippa, M. Recalcati, M. Rabuffetti, M. Ferrarin, A. Casiraghi, Concepts of Motor Learning Applied to a Rehabilitation Protocol Using Biofeedback to Improve Gait in a Chronic Stroke Patient: An A-B System Study With Multiple Gait Analyses, *Neurorehabil. Neural Repair.* 21 (2007) 190–194. doi:10.1177/1545968306290823.
 - [S86] K. Jung, Y. Kim, Y. Cha, T.-S. In, Y.-G. Hur, Y. Chung, Effects of gait training with a cane and an augmented pressure sensor for enhancement of weight bearing over the affected lower limb in patients with stroke: a randomized controlled pilot study., *Clin. Rehabil.* 29 (2015) 135–42. doi:10.1177/0269215514540923.
 - [S87] M. Kassover, C. Tauber, J. Au, J. Pugh, Auditory biofeedback in spastic diplegia, *J. Orthop. Res.* 4 (1986) 246–249.
 - [S88] B. Kaur, N. Manikandan, Effect of auditory feedback on lower limb weight bearing symmetry and gait parameters in patients with hemiparesis, *Indian J. Physiother. Occup. Ther.* 3 (2009) 1–5.
 - [S89] A.H. Khandoker, W.A. Sparrow, R.K. Begg, Tone entropy analysis of augmented information effects on toe-ground clearance when walking, *IEEE Trans. Neural Syst. Rehabil. Eng.* 24 (2016) 1218–1224. doi:10.1109/TNSRE.2016.2538294.
 - [S90] K.-I. Ki, M.-S. Kim, Y. Moon, J.-D. Choi, Effects of auditory feedback during gait training on hemiplegic patients' weight bearing and dynamic balance ability, *J Phys*

- Ther Sci. 27 (2015) 1267–1269. doi:10.1589/jpts.27.1267.
- [S91] V. Krishnan, I. Khoo, P. Marayong, K. DeMars, J. Cormack, Gait Training in Chronic Stroke Using Walk-Even Feedback Device: A Pilot Study, *Neurosci. J.* 2016 (2016) 1–8. doi:10.1155/2016/6808319.
 - [S92] P. Levinger, D. Zeina, A.K. Teshome, E. Skinner, R. Begg, J.H. Abbott, A real time biofeedback using Kinect and Wii to improve gait for post-total knee replacement rehabilitation: a case study report., *Disabil. Rehabil. Assist. Technol.* 11 (2016) 251–262. doi:10.3109/17483107.2015.1080767.
 - [S93] M.D. Lewek, J. Feasel, E. Wentz, F.P. Brooks, M.C. Whitton, Use of Visual and Proprioceptive Feedback to Improve Gait Speed and Spatiotemporal Symmetry Following Chronic Stroke : A Case Series, *Phys. Ther.* 92 (2012) 748–756. doi:10.2522/ptj.20110206.
 - [S94] S.B. Lim, B.C. Horslen, J.R. Davis, J.H.J. Allum, M.G. Carpenter, Benefits of multi-session balance and gait training with multi-modal biofeedback in healthy older adults, *Gait Posture.* 47 (2016) 10–17. doi:10.1016/j.gaitpost.2016.03.017.
 - [S95] C. Ma, D. Wong, W. Lam, A. Wan, W. Lee, Balance Improvement Effects of Biofeedback Systems with State-of-the-Art Wearable Sensors: A Systematic Review, *Sensors.* 16 (2016) 434. doi:10.3390/s16040434.
 - [S96] A.R. Mandel, J.R. Nymark, S.J. Balmer, D.M. Grinnell, M.D. O’Riain, Electromyographic versus rhythmic positional biofeedback in computerized gait retraining with stroke patients, *Arch. Phys. Med. Rehabil.* 71 (1990) 649–654.
 - [S97] F. Massaad, T.M. Lejeune, C. Detrembleur, Reducing the energy cost of hemiparetic gait using center of mass feedback: a pilot study., *Neurorehabil. Neural Repair.* 24 (2010) 338–347. doi:10.1177/1545968309349927.
 - [S98] Z. McKinney, K. Heberer, E. Fowler, M. Greenberg, B. Nowroozi, W. Grundfest, Initial Biomechanical Evaluation of Wearable Tactile Feedback System for Gait Rehabilitation in Peripheral Neuropathy, *Stud. Health Technol. Inform.* (2014) 271–277. doi:10.3233/978-1-61499-375-9-271.
 - [S99] S. Miyazaki, H. Iwakura, Limb-load alarm device for partial-weight-bearing walking exercise., *Med. Biol. Eng. Comput.* 16 (1978) 500–506. doi:10.1007/BF02457799.
 - [S100] R. Montoya, P. Dupui, B. Pagès, P. Bessou, Step-length biofeedback device for walk rehabilitation, *Med. Biol. Eng. Comput.* 32 (1994) 416–420.
 - [S101] M.E. Morris, T.A. Matyas, T.M. Bach, P.A. Goldie, Electrogoniometric feedback: its effect on genu recurvatum in stroke., *Arch. Phys. Med. Rehabil.* 73 (1992) 1147–1154.
 - [S102] W. Nanhoe-Mahabier, J.H. Allum, E.P. Pasman, S. Overeem, B.R. Bloem, The effects of vibrotactile biofeedback training on trunk sway in Parkinson’s disease patients, *Park. Relat. Disord.* 18 (2012) 1017–1021. doi:10.1016/j.parkreldis.2012.05.018.
 - [S103] B. Noehren, J. Scholz, I. Davis, The effect of real-time gait retraining on hip kinematics, pain and function in subjects with patellofemoral pain syndrome., *Br. J. Sports Med.* 45 (2011) 691–696. doi:10.1136/bjsm.2009.069112.
 - [S104] S.J. Olney, G.R. Colborne, C.S. Martin, Joint Angle Feedback and Biomechanical Gait Analysis in Stroke Patients : A Case Report, *Phys. Ther.* 69 (1989) 863–870.
 - [S105] I.L.B. Oude Lansink, L. van Kouwenhove, P.U. Dijkstra, K. Postema, J.M. Hijmans, Effects of interventions on normalizing step width during self-paced dual-belt treadmill walking with virtual reality, a randomised controlled trial, *Gait Posture.* 58 (2017) 121–125. doi:10.1016/j.gaitpost.2017.07.040.
 - [S106] D. Owaki, Y. Sekiguchi, K. Honda, A. Ishiguro, S.I. Izumi, Short-term effect of prosthesis transforming sensory modalities on walking in stroke patients with

- hemiparesis, *Neural Plast.* 2016 (2016) 9 pages. doi:10.1155/2016/6809879.
- [S107] Z. Pataky, D. De León Rodríguez, L. Allet, A. Golay, M. Assal, J.-P. Assal, C.-A. Hauert, Biofeedback for foot offloading in diabetic patients with peripheral neuropathy, *Diabet. Med.* 27 (2010) 61–64. doi:10.1111/j.1464-5491.2009.02875.x.
 - [S108] Z. Pataky, D. De León Rodríguez, A. Golay, M. Assal, J.-P. Assal, C.-A. Hauert, Biofeedback Training for Partial Weight Bearing in Patients After Total Hip Arthroplasty, *Arch. Phys. Med. Rehabil.* 90 (2009) 1435–1438. doi:10.1016/j.apmr.2009.02.011.
 - [S109] O. Pelykh, A.-M. Klein, I. Feist-Pagenstert, C. Schlick, J. Ilmberger, Treatment outcome of visual feedback training in an adult patient with habitual toe walking, *Eur. J. Phys. Rehabil. Med.* (2014).
 - [S110] J.S. Petrofsky, Microprocessor-based gait analysis system to retrain Trendelenburg gait, *Med. Biol. Eng. Comput.* 39 (2001) 140–143. doi:10.1007/BF02345278.
 - [S111] J.S. Petrofsky, The use of electromyogram biofeedback to reduce Trendelenburg gait, *Eur. J. Appl. Physiol.* 85 (2001) 491–495. doi:10.1007/s004210100466.
 - [S112] C. Pizzolato, M. Reggiani, D.J. Saxby, E. Ceseracciu, L. Modenese, D.G. Lloyd, Biofeedback for Gait Retraining Based on Real-Time Estimation of Tibiofemoral Joint Contact Forces, *IEEE Trans. Neural Syst. Rehabil. Eng. A Publ. IEEE Eng. Med. Biol. Soc.* 25 (2017) 1612–1621. doi:10.1109/TNSRE.2017.2683488.
 - [S113] A. Plauche, D. Villarreal, R.D. Gregg, A Haptic Feedback System for Phase-Based Sensory Restoration in Above-Knee Prosthetic Leg Users, *IEEE Trans. Haptics.* 9 (2016) 421–426. doi:10.1109/TOH.2016.2580507.
 - [S114] F. Pu, W. Ren, X. Fan, W. Chen, S. Li, D. Li, Y. Wang, Y. Fan, Real-time feedback of dynamic foot pressure index for gait training of toe-walking children with spastic diplegia., *Disabil. Rehabil.* (2016) 1–5. doi:10.1080/09638288.2016.1212114.
 - [S115] R. Richards, J. Van Den Noort, M. Booij, J. Harlaar, Real-time feedback for gait retraining in knee osteoarthritis: Responses to different types of feedback and instructions, *Gait Posture.* 49S (2016) 70. doi:10.1016/j.gaitpost.2016.07.131.
 - [S116] R.E. Richards, J.C. van den Noort, M. van der Esch, M.J. Booij, J. Harlaar, Effect of real-time biofeedback on peak knee adduction moment in patients with medial knee osteoarthritis: Is direct feedback effective?, *Clin. Biomech.* (2017). doi:10.1016/j.clinbiomech.2017.07.004.
 - [S117] J.A. Sabolich, G.M. Ortega, Sense of Feel for Lower-Limb Amputees: A Phase-One Study, *JPO J. Prosthetics Orthot.* 6 (1994) 36–41.
 - [S118] C.M. Said, H. Nagano, L. James, P. Levinger, M. Taylor, A. Moss, M. Galea, R. Begg, Feasibility of biofeedback gait training to increase minimal toe-clearance in people with stroke, *Int. J. Stroke.* 10 (2015) 77.
 - [S119] M. Schauer, K.-H. Mauritz, Musical motor feedback (MMF) in walking hemiparetic stroke patients: randomized trials of gait improvement., *Clin. Rehabil.* 17 (2003) 713–722. doi:10.1191/0269215503cr668oa.
 - [S120] C. Schenck, T.M. Kesar, Effects of unilateral real-time biofeedback on propulsive forces during gait, *J. Neuroeng. Rehabil.* 14 (2017) 1–10. doi:10.1186/s12984-017-0252-z.
 - [S121] D. Schliessmann, C. Schuld, M. Schneiders, S. Derlien, M. Glöckner, T. Gladow, N. Weidner, R. Rupp, Feasibility of visual instrumented movement feedback therapy in individuals with motor incomplete spinal cord injury walking on a treadmill., *Front. Hum. Neurosci.* 8 (2014) Article 416. doi:10.3389/fnhum.2014.00416.
 - [S122] N.A. Segal, N.A. Glass, P. Teran-Yengle, B. Singh, R.B. Wallace, H.J. Yack, Intensive Gait Training for Older Adults with Symptomatic Knee Osteoarthritis, *Am. J. Phys. Med. Rehabil.* 94 (2015) 848–858.

doi:10.1097/PHM.0000000000000264.

- [S123] J. Shin, Y. Chung, Influence of visual feedback and rhythmic auditory cue on walking of chronic stroke patient induced by treadmill walking in real-time basis, *NeuroRehabilitation*. 41 (2017) 445–452. doi:10.3233/NRE-162139.
- [S124] P.B. Shull, K.L. Lurie, M.R. Cutkosky, T.F. Besier, Training multi-parameter gaits to reduce the knee adduction moment with data-driven models and haptic feedback, *J. Biomech.* 44 (2011) 1605–1609. doi:10.1016/j.jbiomech.2011.03.016.
- [S125] K.H. Sienko, M.D. Balkwill, L.I.E. Oddsson, C. Wall, The effect of vibrotactile feedback on postural sway during locomotor activities., *J. Neuroeng. Rehabil.* 10 (2013) 93–98. doi:10.1186/1743-0003-10-93.
- [S126] M. Simic, K.L. Bennell, M.A. Hunt, T. V. Wrigley, R.S. Hinman, Contralateral cane use and knee joint load in people with medial knee osteoarthritis: The effect of varying body weight support, *Osteoarthr. Cartil.* 19 (2011) 1330–1337. doi:10.1016/j.joca.2011.08.008.
- [S127] M. Simic, M.A. Hunt, K.L. Bennell, R.S. Hinman, T. V. Wrigley, Trunk lean gait modification and knee joint load in people with medial knee osteoarthritis: The effect of varying trunk lean angles, *Arthritis Care Res.* 64 (2012) 1545–1553. doi:10.1002/acr.21724.
- [S128] J.J. Tate, C.E. Milner, Sound-Intensity Feedback During Running Reduces Loading Rates and Impact Peak, *J. Orthop. Sport. Phys. Ther.* 47 (2017) 565–569. doi:10.2519/jospt.2017.7275.
- [S129] P. Teran-Yengle, R. Birkhofer, M.A. Weber, K. Patton, E. Thatcher, H.J. Yack, Efficacy of Gait Training With Real-Time Biofeedback in Correcting Knee Hyperextension Patterns in Young Women, *J. Orthop. Sports Phys. Ther.* 41 (2011) 948–952. doi:10.2519/jospt.2011.3660.
- [S130] P. Teran-Yengle, K.J. Cole, H.J. Yack, Short and long-term effects of gait retraining using real-time biofeedback to reduce knee hyperextension pattern in young women, *Gait Posture*. 50 (2016) 185–189. doi:10.1016/j.gaitpost.2016.08.019.
- [S131] O. Tirosh, A. Cambell, R.K. Begg, W.A. Sparrow, Biofeedback training effects on minimum toe clearance variability during treadmill walking, *Ann. Biomed. Eng.* 41 (2013) 1661–1669. doi:10.1007/s10439-012-0673-6.
- [S132] S.D. Uhlrich, A. Silder, G.S. Beaupre, P.B. Shull, S.L. Delp, Subject-specific toe-in or toe-out gait modifications reduce the larger knee adduction moment peak more than a non-personalized approach, *J. Biomech.* (2017). doi:10.1016/j.jbiomech.2017.11.003.
- [S133] J.C. van den Noort, F. Steenbrink, S. Roeles, J. Harlaar, Real-time visual feedback for gait retraining: toward application in knee osteoarthritis, *Med. Biol. Eng. Comput.* 53 (2015) 275–286. doi:10.1007/s11517-014-1233-z.
- [S134] M. van der Krogt, L. van Gelder, I. van de Port, A. Buizer, J. Harlaar, Real-time feedback to improve gait in children with cerebral palsy, *Gait Posture*. 42 (2015) S83–S84. doi:10.1016/j.gaitpost.2015.06.153.
- [S135] R.P. Van Der Logt, O. Findling, H. Rust, O. Yaldizli, J.H.J. Allum, The effect of vibrotactile biofeedback of trunk sway on balance control in multiple sclerosis, *Mult. Scler. Relat. Disord.* 8 (2016) 58–63. doi:10.1016/j.msard.2016.05.003.
- [S136] L. van Gelder, A.T.C. Booth, I. van de Port, A.I. Buizer, J. Harlaar, M.M. van der Krogt, Real-time feedback to improve gait in children with cerebral palsy, *Gait Posture*. 52 (2017) 76–82. doi:10.1016/j.gaitpost.2016.11.021.
- [S137] L.L. Verhoeff, C.G.C. Horlings, L.J.F. Janssen, S.A. Bridenbaugh, J.H.J. Allum, Effects of biofeedback on trunk sway during dual tasking in the healthy young and elderly, *Gait Posture*. 30 (2009) 76–81. doi:10.1016/j.gaitpost.2009.03.002.

- [S138] S.C. Walker, P.A. Helm, L.A. Lavery, Gait pattern alteration by functional sensory substitution in healthy subjects and in diabetic subjects with peripheral neuropathy, *Arch. Phys. Med. Rehabil.* 78 (1997) 853–856. doi:10.1016/S0003-9993(97)90199-4.
- [S139] J.W. Wheeler, P.B. Shull, T.F. Besier, Real-time knee adduction moment feedback for gait retraining through visual and tactile displays., *J. Biomech. Eng.* 133 (2011) Article 041007. doi:10.1115/1.4003621.
- [S140] S.C. White, R.M. Lifeso, Altering asymmetric limb loading after hip arthroplasty using real-time dynamic feedback when walking, *Arch. Phys. Med. Rehabil.* 86 (2005) 1958–1963. doi:10.1016/j.apmr.2005.04.010.
- [S141] K.N. Winfree, I. Pretzer-Aboff, D. Hilgart, R. Aggarwal, M. Behari, S. Agrawal, An untethered shoe with vibratory feedback for improving gait of Parkinson's Patients: The PDShoe, *Proc. Annu. Int. Conf. IEEE Eng. Med. Biol. Soc. EMBS.* (2012) 1202–1205. doi:10.1109/EMBC.2012.6346152.
- [S142] K.N. Winfree, I. Pretzer-Aboff, D. Hilgart, R. Aggarwal, M. Behari, S.K. Agrawal, The effect of step-synchronized vibration on patients with parkinson's disease: Case studies on subjects with freezing of gait or an implanted deep brain stimulator, *IEEE Trans. Neural Syst. Rehabil. Eng.* 21 (2013) 806–811. doi:10.1109/TNSRE.2013.2250308.
- [S143] C.M. Wood, K. Kipp, Use of audio biofeedback to reduce tibial impact accelerations during running, *J. Biomech.* 47 (2014) 1739–1741. doi:10.1016/j.jbiomech.2014.03.008.
- [S144] A. Woolley-Hart, I. Musa, E. Rodger, The use of electromyographic biofeedback in the retraining of gait following stroke, *Chest, Hear. Stroke J.* 2 (1977) 3–9.
- [S145] L. Yang, P.S. Dyer, R.J. Carson, J.B. Webster, K. Bo Foreman, S.J.M. Bamberg, Utilization of a lower extremity ambulatory feedback system to reduce gait asymmetry in transtibial amputation gait, *Gait Posture.* 36 (2012) 631–634. doi:10.1016/j.gaitpost.2012.04.004.
- [S146] J. Xu, T. Bao, U.H. Lee, C. Kinnaird, W. Carender, Y. Huang, K.H. Sienko, P.B. Shull, Configurable, wearable sensing and vibrotactile feedback system for real-time postural balance and gait training: Proof-of-concept, *J. Neuroeng. Rehabil.* 14 (2017) 1–10. doi:10.1186/s12984-017-0313-3.
- [S147] S.C. Yen, J.M. Landry, M. Wu, Augmented multisensory feedback enhances locomotor adaptation in humans with incomplete spinal cord injury, *Hum. Mov. Sci.* 35 (2014) 80–93. doi:10.1016/j.humov.2014.03.006.
- [S148] W.-G. Yoo, Effect of a Portable EMG-based Combined Biofeedback Device (PECBD) for the Rectus Femoris, Biceps Femoris, and Tibialis Anterior Muscles on Stroke Gait., *J. Phys. Ther. Sci.* 24 (2012) 1229–1232. doi:10.1589/jpts.24.1229.