



Contents lists available at [ScienceDirect](#)

Methods

journal homepage: www.elsevier.com/locate/ymeth



Corrigendum

Corrigendum to 'Methods to account for movement and flexibility in cryo-EM data processing' [Methods 100 (2016) 35–41]



S. Rawson ^a, M.G. Iadanza ^b, N.A. Ranson ^b, S.P. Muench ^{a,*}

^a School of Biomedical Sciences, Faculty of Biological Sciences, University of Leeds, Leeds LS2 9JT, UK

^b School of Molecular and Cellular Biology, Faculty of Biological Sciences, University of Leeds, Leeds LS2 9JT, UK

The authors regret that after communication with J. Rubinstein we have re-run alignparts_lmbfgs using a higher lambda value (1d12) than that in the example given in the protocol to take account of the larger image pixel values with a FEI Falcon camera compared to a Gatan K2 summit camera. This has resulted in a significantly improved performance of the program.

The authors would like to apologise for any inconvenience caused.

DOI of original article: <http://dx.doi.org/10.1016/j.ymeth.2016.03.011>

* Corresponding author.

E-mail address: s.p.muench@leeds.ac.uk (S.P. Muench).

<http://dx.doi.org/10.1016/j.ymeth.2016.05.006>

1046-2023/© 2016 The Authors. Published by Elsevier Inc.

This is an open access article under the CC BY license (<http://creativecommons.org/licenses/by/4.0/>).