



UNIVERSITY OF LEEDS

This is a repository copy of *Precision of maxillo-mandibular registration with intraoral scanners in vitro*.

White Rose Research Online URL for this paper:  
<http://eprints.whiterose.ac.uk/152693/>

Version: Accepted Version

---

**Article:**

Gintaute, A, Keeling, AJ [orcid.org/0000-0003-4598-3744](https://orcid.org/0000-0003-4598-3744), Osnes, CA  
[orcid.org/0000-0003-4652-3854](https://orcid.org/0000-0003-4652-3854) et al. (3 more authors) (2020) Precision of  
maxillo-mandibular registration with intraoral scanners in vitro. *Journal of Prosthodontic  
Research*, 64 (2). pp. 114-119. ISSN 1883-1958

<https://doi.org/10.1016/j.jpor.2019.05.006>

---

© 2019 Japan Prosthodontic Society. Published by Elsevier Ltd. All rights reserved. This  
manuscript version is made available under the CC-BY-NC-ND 4.0 license  
<http://creativecommons.org/licenses/by-nc-nd/4.0/>.

**Reuse**

This article is distributed under the terms of the Creative Commons Attribution-NonCommercial-NoDerivs  
(CC BY-NC-ND) licence. This licence only allows you to download this work and share it with others as long  
as you credit the authors, but you can't change the article in any way or use it commercially. More  
information and the full terms of the licence here: <https://creativecommons.org/licenses/>

**Takedown**

If you consider content in White Rose Research Online to be in breach of UK law, please notify us by  
emailing [eprints@whiterose.ac.uk](mailto:eprints@whiterose.ac.uk) including the URL of the record and the reason for the withdrawal request.



[eprints@whiterose.ac.uk](mailto:eprints@whiterose.ac.uk)  
<https://eprints.whiterose.ac.uk/>

**Table 2.**

Point-to-point distances (means, in mm with 95% CI) for groups and scanners. The p-value represents the ANOVA results for each group across scanners.

Distance	Group	Point-to-point distance (95% CI)			p-value
		CEREC	TRIOS	PLANMECA	
UR7-LR7	A	1.03 (0.93–1.13)	0.94 (0.80–1.08)	0.80 (0.61–0.99)	0.023
	B	0.81 (0.75-0.87)	0.65 (0.49-0.81)	0.55 (0.30-0.80)	0.032
	C	0.86 (0.67-1.03)	0.68 (0.52-0.84)	0.85 (0.50-1.20)	0.29
	D	0.90 (0.85-0.95)	0.77 (0.68-0.86)	0.98 (0.60-1.36)	0.283
UL7-LL7	A	0.52 (0.41-0.63)	0.37 (0.30-0.44)	0.36 (0.17-0.55)	0.062
	B	0.72 (0.66-0.78)	0.52 (0.40-0.64)	0.41 (0.26-0.56)	0.001
	C	0.72 (0.62-0.82)	0.37 (0.32-0.42)	0.55 (0.10-1.00)	0.076
UR3-LR3	A	2.17 (2.15-2.19)	2.21 (2.12-2.30)	2.28 (2.17-2.39)	0.063
	B	2.14 (2.08-2.20)	2.08 (2.01-2.15)	2.19 (2.04-2.34)	0.205
	C	2.13 (2.04-2.22)	2.15 (2.10-2.20)	2.14 (2.05-2.23)	0.913
	D	1.85 (1.81-1.89)	1.83 (1.81-1.85)	2.02 (1.81-2.23)	0.035
	E	2.41 (2.37-2.45)	2.45 (2.38-2.52)	2.38 (2.28-2.48)	0.213
	F	1.94 (1.92-1.96)	1.93 (1.88-1.98)	1.63 (1.57-1.69)	< 0.001
UL3-LL3	A	0.84 (0.82-0.86)	0.83 (0.76-0.90)	0.82 (0.70-0.94)	0.863
	B	0.79 (0.75-0.83)	0.77 (0.66-0.88)	0.80 (0.64-0.96)	0.884
	C	0.74 (0.72-0.76)	0.78 (0.71-0.85)	0.82 (0.72-0.92)	0.177
	D	0.50 (0.46-0.54)	0.44 (0.38-0.50)	0.37 (0.31-0.43)	0.005
	E	1.08 (1.06-1.10)	1.10 (1.05-1.15)	0.92 (0.86-0.98)	< 0.001
	F	0.66 (0.61-0.71)	0.68 (0.63-0.73)	0.46 (0.37-0.55)	< 0.001