

This is a repository copy of *Nanophotonic platform for the highly sensitive detection of trypsin enzyme in human urine*.

White Rose Research Online URL for this paper:

<https://eprints.whiterose.ac.uk/id/eprint/227294/>

Version: Published Version

Article:

Bakshi, Shrishty, Li, Kezheng, Sahoo, Pankaj K. et al. (1 more author) (2025)

Nanophotonic platform for the highly sensitive detection of trypsin enzyme in human urine. Chemical Communications. pp. 5321-5324. ISSN 1364-548X

<https://doi.org/10.1039/d4cc04840g>

Reuse

This article is distributed under the terms of the Creative Commons Attribution (CC BY) licence. This licence allows you to distribute, remix, tweak, and build upon the work, even commercially, as long as you credit the authors for the original work. More information and the full terms of the licence here:

<https://creativecommons.org/licenses/>

Takedown

If you consider content in White Rose Research Online to be in breach of UK law, please notify us by emailing eprints@whiterose.ac.uk including the URL of the record and the reason for the withdrawal request.



Cite this: *Chem. Commun.*, 2025, 61, 5321

Received 18th September 2024,
Accepted 12th February 2025

DOI: 10.1039/d4cc04840g

rsc.li/chemcomm

Nanophotonic platform for the highly sensitive detection of trypsin enzyme in human urine†

Shrishty Bakshi, * Kezheng Li,  Pankaj K. Sahoo and Thomas F. Krauss

Trypsin is an important indicator of pancreatic disease. We introduce a nanophotonic test for the detection of trypsin. Our approach is based on the hydrolysis of β -casein by trypsin. We demonstrate 0.1 ng mL⁻¹ of trypsin detection in human urine and across the clinically relevant range.

Pancreatic diseases are some of the most common causes of illness around the globe, affecting a significant number of people globally.^{1,2} The incidence of pancreatitis, one of the most common pancreas diseases, results in 12 000 hospital admissions per year in the UK alone³ and the problem is rapidly increasing worldwide.⁴ These observations call for a better, highly sensitive testing technology that is easily accessible and that can help detect pancreatic diseases at an early stage for better management and timely treatment. Trypsin is an important enzyme, used for studying pancreatic diseases.^{5,6} Trypsin is involved in the hydrolysis of protein molecules into shorter peptide fragments to enhance their absorption.^{7,8} Studies have found that there is no trypsin found in the urine of healthy people. However, the urine of patients who have undergone pancreas transplants contains high levels of trypsin, *i.e.* concentrations on the order of 10's of $\mu\text{g mL}^{-1}$.^{9,10} We also note that trypsin plays an important role in the progression of pancreatic cancer and development of diseases such as cystic fibrosis, viral diseases, *etc.*^{11–13}

The techniques commonly used for trypsin detection are enzyme-linked immunosorbent assay (ELISA),¹⁴ mass spectrometry,¹⁵ fluorescent-based assays,⁹ radioimmunoassay¹⁶ and gel electrophoresis.¹⁷ The major drawback of these methods is the requirement for labelling and the need for bulky and sensitive instrumentation as well as skilled labour, which results in high costs and lengthy analysis times, thus reducing the uptake of such tests. Efforts to develop point-of-care

biosensing devices for trypsin detection are underway,¹⁸ but most approaches are based on fluorescence,^{11,19–21} which requires labelling and more complicated read-out instruments. Alternatively, tests that are more suitable for near-patient use tend to have lower sensitivity and a limited dynamic range.²² The ESI,† Table S1 provides a more detailed comparison of different label-free sensing technologies developed for the detection of trypsin. Overall, we observe that a highly sensitive trypsin biosensor with a large dynamic range that is easy to use and can be realised cost-effectively has not yet been demonstrated.

Here, we introduce such a biosensor. Our modality is based on a wavelength-scale grating that supports guided mode resonances (GMRs), which can be readily excited with collimated light from an LED. This modality is much easier to implement than many other photonic sensors that require the difficult coupling of laser light into a waveguide of submicron cross-section. By continuously varying (“chirping”) the period of the grating, we translate spectral information into spatial information such that the readout can be performed with a simple CMOS camera, thereby removing the requirement for a high-resolution spectrometer.²³ In addition, to address any drift caused by the light source or temperature-related fluctuations, a reference channel is placed in close proximity to the measurement channel. We have also shown that the sensor can be realised in a handheld format without loss of performance²⁴ and that it is robust against thermal fluctuations and mechanical vibration;²⁵ hence, it is ideally suited for near-patient deployment. Here, we present evidence that this modality is also effective for monitoring trypsin levels in clinical fluids, especially human urine. The sensing principle is based on the hydrolysis of a β -casein layer immobilised on the sensor surface by the trypsin enzyme.

A robust and stable surface chemistry ensures the sensitive detection of analytes of interest from a complex matrix. Fig. 1(a–d) describes the preparation and operation of the biological interface used here. The interface employs polydopamine surface chemistry, which we have shown to be stable and effective.^{26,27} The polydopamine layer is applied to the grating to immobilise the β -casein substrate. The grating is

School of Physics, Engineering and Technology, University of York, York, YO10 5DD, UK. E-mail: shrishty.bakshi@york.ac.uk

† Electronic supplementary information (ESI) available. See DOI: <https://doi.org/10.1039/d4cc04840g>






Fig. 1 Structure of the sensor and biological interface for the detection of trypsin: (a) sketch of the silicon nitride-based grating sensor; (b) coating of the sensor surface with polydopamine surface chemistry; (c) in-flow immobilisation of β -casein on the coated sensor surface; (d) sensor surface after the action of the enzyme.

fabricated using electron beam lithography, and it exhibits a resonance when illuminated by light at 650 nm wavelength (Fig. 2(a)). As we have already shown, the sensors can also be fabricated using nanoimprint lithography for larger-scale commercial purposes.²⁴

We first perform the measurements in phosphate buffer saline (PBS) solution. Typically, GMR sensors exhibit an increase in resonance wavelength (red shift) when proteins bind to the sensor surface. This occurs because the bound material increases the refractive index of the cladding, thereby causing the resonant mode to shift. Since we chirp the grating's period and illuminate with a fixed wavelength, this will translate the resonant position towards a lower period as the refractive index increases. The position of the resonance is then captured by a CMOS camera (Fig. 2a). In the case of detecting enzymatic activity, the resonance moves in the opposite direction (blue-shift), because trypsin digests the β -casein immobilised on the surface, resulting in a mass loss on the sensor surface. By tracking the resonance position over time, the kinetics of the digestion of the β -casein layer is monitored. Fig. 3a shows the sequence of the full assay, starting from the immobilisation of β -casein, resulting in mass gain, to washing,




Fig. 2 (a) Optical micrograph of two chirped gratings exhibiting resonances in reflection upon illumination with 650 nm wavelength light from a halogen source. We use a monochromator to reduce the linewidth to $\Delta\lambda < 1$ nm; (b) scanning electron microscope image of the grating.




Fig. 3 (a) Full assay for the detection of trypsin (10 ng mL^{-1}) in PBS; (b) dynamic range graph for concentrations from 0.1 ng mL^{-1} to 1000 ng mL^{-1} . The inset shows the unbinding curves for the different trypsin concentrations between 0.1 ng mL^{-1} – 1000 ng mL^{-1} as the β -casein is digested by the enzyme.

to the introduction of trypsin (10 ng mL^{-1}), resulting in mass loss. Fig. 3b then shows the dynamic range for different trypsin concentrations, varying between 0.1 ng mL^{-1} and 1000 ng mL^{-1} . We note that for a concentration of 0.1 ng mL^{-1} , the observed shift is approx. 0.066 nm , which is well above the 3σ (mean $3\sigma = 0.019 \text{ nm} \pm 0.01 \text{ nm}$ ($n = 31 \pm \text{SD}$)) value of the system. For the higher concentrations of up to 1000 ng mL^{-1} , we observe a linear trend with an R^2 value of 0.99383 ; this high R^2 value highlights that our sensor is suitable for quantifying different trypsin concentrations. The inset in Fig. 3b displays the individual unbinding curves for various concentrations. To demonstrate the specificity of the sensor against interfering molecules in human urine, we also studied the effect of urea and glucose. For reference, when using 100 ng mL^{-1} of trypsin in PBS, we observe a resonance shift of $0.228 \pm 0.02 \text{ nm}$ ($n = 3$). When adding 2% urea, the shift is $0.231 \pm 0.01 \text{ nm}$ ($n = 3$) for 100 ng mL^{-1} trypsin concentration and when adding $100 \mu\text{g mL}^{-1}$ of glucose, the shift is $0.232 \pm 0.03 \text{ nm}$ ($n = 3$) for 100 ng mL^{-1} trypsin concentration. Given that all values are comparable we can be confident that the main interferences in human urine only have a negligible effect on the sensor readout.

Finally, we assess the suitability of our sensor to perform trypsin detection in a real human matrix, here undiluted



human urine. As before, we show the sequence of the assay for enzyme detection in human urine (Fig. 4a). The first step of β -casein immobilisation is the same, as shown in Fig. 3a(i). The second step, urine introduction, causes an upward shift because urine has a slightly higher refractive index than PBS (Fig. 4a(ii)). We note that the resonance curve is stable, indicating that there is neither fouling nor any other enzymatic reaction from interfering molecules (note that the urine was pooled from a sample of healthy volunteers, which may contain a variety of different molecules), which gives us confidence that our assay is specific to trypsin. When we introduce the same urine spiked with trypsin (Fig. 4a(iii)), we see a clear unbinding curve, indicating that the trypsin digests the substrate.

We used the same concentration range as before, 0.1 ng mL^{-1} to 1000 ng mL^{-1} , and Fig. 4b shows the corresponding resonance shifts. The shift observed for 0.1 ng mL^{-1} is approximately 0.065 nm , which is above the 3 Sigma value. As before, we observe a linear trend with a high R^2 value of 0.99469 .

In conclusion, we have demonstrated a diagnostic test based on guided mode resonances that is able to detect the trypsin

enzyme at concentrations as low as 0.1 ng mL^{-1} . The response of the test is linear over a concentration range spanning 4 orders of magnitude, which includes the clinically relevant range for the diagnosis of pancreatic disease. The test performs equally well in phosphate buffered saline (PBS) as in undiluted human urine, which gives us confidence in its robustness and its ability to cope with the large variety of patient samples that are encountered in practice. The test can be used for the screening and monitoring at the point of care as we have previously demonstrated its realisation as a handheld instrument.²⁴ In addition to the specific pancreatic disease application we describe here, the work opens up the possibility of detecting other types of enzymes with the choice of appropriate substrates. By combining the high performance of laboratory instruments with the convenience of a near-patient test, we aim to make this important diagnostic more widely available to improve the early detection and improved management of disease. In this scenario, we note that long-term storage of the functionalised sensors is required, which can be achieved by stabilising the sensor surface with sugars and polyhydric alcohols.^{28,29}

We acknowledge funding by UK Research & Innovation under grants EP/V047434/1, EP/X037770/1 and EP/V047663/1.

Data availability

Data will be made available on reasonable request.

Conflicts of interest

The authors declare there are no conflicts of interest.

Notes and references

- 1 J. Huang, V. Lok, C. H. Ngai, L. Zhang, J. Yuan, X. Q. Lao, K. Ng, C. Chong, Z. J. Zheng and M. C. S. Wong, *Gastroenterology*, 2021, **160**, 744–754.
- 2 A. Pourshams, *et al.*, *Lancet Gastroenterol. Hepatol.*, 2019, **4**, 934–947.
- 3 T. C. Hall, G. Garcea, M. A. Webb, D. Al-Leswas, M. S. Metcalfe and A. R. Dennison, *J. Eval. Clin. Pract.*, 2014, **20**, 203–207.
- 4 G. Ouyang, G. Pan, Q. Liu, Y. Wu, Z. Liu, W. Lu, S. Li, Z. Zhou and Y. Wen, *BMC Med.*, 2020, **18**, 388.
- 5 J. Yin, T. Cui, Y. Yang and T.-L. Ren, *Chemosensors*, 2023, **11**, 469.
- 6 V. V. Gummadi and T. Gonska, *Pancreas*, 2024, 10–97.
- 7 J. Kaur and P. K. Singh, *Crit. Rev. Anal. Chem.*, 2022, **52**, 949–967.
- 8 V. G. Vertiprakhov and N. V. Ovchinnikova, *Front. Physiol.*, 2022, **13**, 874664.
- 9 J. Xu, E. Prost, K. Haupt and B. Tse Sum Bui, *Sens. Actuators, B*, 2018, **258**, 10–17.
- 10 W. A. See and J. L. Smith, *Transplantation*, 1991, **52**, 630–633.
- 11 C. Munoz-San Martin, M. Pedrero, M. Gamella, A. Montero-Calle, R. Barderas, S. Campuzano and J. M. Pingarron, *Anal. Bioanal. Chem.*, 2020, **412**, 6177–6188.
- 12 P. Nyberg, M. Ylipalosaari, T. Sorsa and T. Salo, *Exp. Cell Res.*, 2006, **312**, 1219–1228.
- 13 O. G. Famutimi, V. G. Adebisi, B. G. Akinmolu, O. V. Dada and I. O. Adewale, *Futur. J. Pharm. Sci.*, 2024, **10**, 126.
- 14 M. A. Seia, P. W. Stege, S. V. Pereira, I. E. De Vito, J. Raba and G. A. Messina, *Anal. Biochem.*, 2014, **463**, 31–37.
- 15 P. A. Everley, C. A. Gartner, W. Haas, A. Saghatelian, J. E. Elias, B. F. Cravatt, B. R. Zetter and S. P. Gygi, *Mol. Cell. Proteomics*, 2007, **6**, 1771–1777.




Fig. 4 (a) Detection of trypsin (0.1 ng mL^{-1}) in undiluted human urine; (b) dynamic range data for the detection of different trypsin concentrations in undiluted human urine, ranging from 0.1 ng mL^{-1} to 1000 ng mL^{-1} . The inset shows the unbinding curves of β -casein when exposed to different trypsin concentrations spiked in human urine between 0.1 ng mL^{-1} and 1000 ng mL^{-1} .



- 16 C. O'Connor, M. O'Donnell and K. McGeeney, *Clin. Chim. Acta*, 1981, **114**, 29–35.
- 17 R. B. Lefkowitz, G. W. Schmid-Schonbein and M. J. Heller, *Electrophoresis*, 2010, **31**, 2442–2451.
- 18 G. Erturk, M. Hedstrom and B. Mattiasson, *Biosens. Bioelectron.*, 2016, **86**, 557–565.
- 19 S. Xu, F. Zhang, L. Xu, X. Liu, P. Ma, Y. Sun, X. Wang and D. Song, *Sens. Actuators, B*, 2018, **273**, 1015–1021.
- 20 L. Hu, S. Han, S. Parveen, Y. Yuan, L. Zhang and G. Xu, *Biosens. Bioelectron.*, 2012, **32**, 297–299.
- 21 M. Wang, D. Su, G. Wang and X. Su, *Anal. Bioanal. Chem.*, 2018, **410**, 6891–6900.
- 22 M. Nemati, A. Santos, T. Kumeria and D. Losic, *Anal. Chem.*, 2015, **87**, 9016–9024.
- 23 G. J. Triggs, Y. Wang, C. P. Reardon, M. Fischer, G. J. O. Evans and T. F. Krauss, *Optica*, 2017, **4**, 229–234.
- 24 A. Drayton, K. Li, M. Simmons, C. Reardon and T. F. Krauss, *Opt. Express*, 2020, **28**, 32239–32248.
- 25 K. Li, N. J. Suliali, P. K. Sahoo, C. D. Silver, M. Davrandi, K. Wright, C. Reardon, S. D. Johnson and T. F. Krauss, *ACS Sens.*, 2024, **9**, 1857–1865.
- 26 S. Bakshi, K. Li, P. Dong, I. Barth, C. Kunstmann-Olsen, S. Johnson and T. F. Krauss, *Talanta*, 2024, **268**, 125300.
- 27 S. Bakshi, P. K. Sahoo, K. Li, S. Johnson, M. J. Raxworthy and T. F. Krauss, *Biosens. Bioelectron.*, 2023, **242**, 115743.
- 28 S. Gerlsma, *J. Biol. Chem.*, 1968, **243**, 957–961.
- 29 J. F. Back, D. Oakenfull and M. B. Smith, *Biochemistry*, 1979, **18**, 5191–5196.

